


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122120\
 Data File : BN013231.D
 Acq On : 21 Dec 2020 08:59
 Operator : CG/JU
 Sample : DFTPP32
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP32

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN120320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Dec 21 10:25:14 2020



AutoFind: Scans 2504, 2505, 2506; Background Corrected with Scan 2497

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	52.5	95312	PASS
68	69	0.00	2	1.7	1509	PASS
69	198	0.00	100	49.6	90077	PASS
70	69	0.00	2	0.5	468	PASS
127	198	10	80	50.7	92200	PASS
197	198	0.00	2	0.4	724	PASS
198	198	100	100	100.0	181680	PASS
199	198	5	9	6.1	11111	PASS
275	198	10	60	29.9	54253	PASS
365	198	1	100	4.7	8505	PASS
441	198	0.01	100	16.2	29504	PASS
442	198	50	100	96.5	175232	PASS
443	442	15	24	19.8	34656	PASS

Scan 2497 (17.175 min): BN013231.D
DFTPP32

m/z	Abundance
40.00	493.0
43.00	361.0
44.00	2423.0
69.10	496.0
73.00	2930.0
79.00	410.0
147.00	1023.0
193.00	552.0
207.00	1914.0
209.00	354.0
221.00	744.0
281.00	796.0
355.10	560.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

m/z	Abundance
37.00	1066.0
38.10	1623.0
39.10	9819.0
40.00	928.0
41.00	595.0
43.10	385.0
44.00	2802.0
45.00	376.0
49.10	1315.0
50.00	36472.0
51.10	120368.0
52.00	5984.0
53.10	524.0
55.00	1072.0
56.00	3605.0
57.00	7192.0
58.10	306.0
61.00	2032.0
62.00	1877.0
63.00	4848.0
63.90	759.0
65.10	2568.0
67.90	2043.0
69.00	116688.0
70.10	634.0
73.00	2536.0
74.00	14117.0
75.00	19496.0
76.10	6921.0
77.10	124472.0
78.00	8403.0
79.10	9305.0
80.00	7396.0
81.00	9793.0
82.00	2470.0
83.00	1995.0
83.90	421.0
84.90	1663.0
86.00	3102.0
87.00	1684.0
88.10	468.0
91.00	1892.0
92.00	2170.0
93.00	17152.0
94.00	1693.0
94.90	411.0
95.30	304.0
96.00	791.0
98.00	14794.0
99.00	10179.0
100.00	953.0
101.00	5044.0
102.00	308.0
103.00	1816.0
104.00	3878.0
105.00	3502.0
106.00	1175.0
107.00	38152.0
108.00	6058.0
109.00	1135.0
110.00	61824.0
111.00	8981.0
112.00	1356.0
112.90	581.0
116.00	2030.0
117.00	45680.0
118.00	3575.0
119.00	519.0
120.00	411.0
122.00	3393.0
123.00	3889.0
124.00	1926.0
124.90	1894.0
126.20	680.0
127.00	114440.0
128.00	8707.0
129.00	46552.0
130.00	4213.0
131.10	825.0
132.30	399.0
133.10	314.0
134.00	1676.0
135.00	4310.0
136.00	1356.0
136.90	1926.0
137.90	495.0
139.90	764.0
141.00	6725.0
141.90	1965.0
142.90	1518.0
144.00	591.0
144.80	589.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

146.00	1302.0
147.00	3474.0
148.00	9693.0
149.10	1673.0
151.00	1014.0
151.80	546.0
152.90	1920.0
154.00	1290.0
155.00	4139.0
156.00	5173.0
156.90	1005.0
158.00	1362.0
158.90	1039.0
160.00	2848.0
161.00	2493.0
162.20	680.0
163.00	410.0
164.00	503.0
165.00	2144.0
166.00	2160.0
167.00	15403.0
168.10	7853.0
169.00	1230.0
170.10	837.0
171.00	994.0
172.00	1654.0
173.00	1422.0
174.00	3229.0
175.00	5731.0
175.90	1810.0
177.00	2180.0
177.90	957.0
179.00	9824.0
180.00	6986.0
181.00	2772.0
182.00	507.0
183.00	311.0
184.10	1113.0
185.10	4849.0
186.10	30072.0
187.00	9436.0
188.00	1182.0
189.00	2740.0
189.90	675.0
191.00	1417.0
192.00	2948.0
193.00	3628.0
194.00	840.0
195.10	452.0
196.00	7876.0
198.00	206400.0
199.00	14033.0
200.00	1197.0
201.00	435.0
201.50	880.0
202.00	387.0
203.00	1649.0
204.00	7886.0
205.00	14671.0
206.00	58616.0
207.10	9317.0
208.00	2068.0
209.10	964.0
210.20	915.0
211.10	2601.0
211.80	355.0
214.90	860.0
216.00	1382.0
217.00	16712.0
218.00	2229.0
221.10	6788.0
222.00	2220.0
223.00	3755.0
224.00	31776.0
225.00	7825.0
226.10	1039.0
227.00	14567.0
228.00	1937.0
229.00	2697.0
230.00	548.0
231.00	1209.0
233.00	319.0
234.00	1094.0
235.00	1461.0
236.00	762.0
237.10	1256.0
238.90	698.0
240.10	449.0
241.00	807.0
242.00	1769.0
243.00	2759.0
244.00	25992.0
245.10	3569.0
246.00	4344.0
247.00	687.0
249.00	1211.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

252.00	339.0
253.10	780.0
255.00	129248.0
256.00	19872.0
257.10	1647.0
258.00	8980.0
259.10	851.0
261.10	389.0
264.20	363.0
265.00	3626.0
266.10	487.0
271.00	567.0
273.00	3993.0
274.00	9825.0
275.00	55216.0
276.00	5373.0
277.00	5152.0
278.10	801.0
281.10	530.0
282.00	404.0
283.00	681.0
284.10	380.0
285.00	951.0
291.00	411.0
292.90	868.0
294.80	537.0
296.00	18376.0
297.00	2523.0
302.30	323.0
303.10	2111.0
304.00	426.0
308.80	339.0
313.90	689.0
315.00	2529.0
316.10	1120.0
321.00	809.0
323.00	5299.0
323.90	735.0
326.90	1348.0
333.00	511.0
334.00	3955.0
335.00	1088.0
340.90	650.0
345.90	1124.0
352.00	1905.0
353.00	987.0
354.00	1546.0
355.00	361.0
365.00	7425.0
366.00	1207.0
371.00	328.0
372.00	1968.0
373.00	391.0
383.10	649.0
392.10	325.0
401.80	621.0
402.90	1039.0
404.00	315.0
420.90	1412.0
422.00	1156.0
423.00	7602.0
423.90	1920.0
441.00	27216.0
442.00	139456.0
443.10	30368.0
444.00	2537.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

m/z	Abundance
36.90	523.0
38.10	2055.0
39.10	8240.0
40.10	537.0
40.90	479.0
43.00	551.0
44.00	3118.0
44.80	347.0
49.00	850.0
50.00	31064.0
51.00	111400.0
52.10	6437.0
53.00	433.0
55.00	1045.0
56.00	3347.0
57.00	6702.0
57.90	392.0
61.00	1194.0
62.00	1755.0
63.00	4279.0
64.00	497.0
65.10	1810.0
66.90	399.0
68.00	1647.0
69.00	105232.0
70.00	437.0
73.00	2427.0
74.00	12408.0
75.00	16310.0
76.10	5714.0
77.00	121352.0
78.00	7768.0
79.00	8633.0
80.00	6816.0
81.00	9340.0
82.00	2248.0
83.00	1892.0
85.00	1623.0
86.10	2971.0
87.00	1225.0
87.90	350.0
90.90	1877.0
92.00	2829.0
93.00	16400.0
94.00	1383.0
95.10	528.0
96.00	755.0
97.10	742.0
98.00	13918.0
99.00	8851.0
99.90	873.0
101.00	4669.0
102.90	1675.0
103.90	3865.0
105.00	3366.0
106.00	942.0
107.00	35312.0
108.00	4880.0
109.00	769.0
110.00	59000.0
111.00	9128.0
112.00	1467.0
112.90	316.0
116.00	2010.0
117.00	41248.0
118.00	2204.0
118.90	529.0
120.00	673.0
122.00	2992.0
123.00	4435.0
124.00	1948.0
124.90	1452.0
127.00	107328.0
128.00	9404.0
129.00	48512.0
130.00	3757.0
131.00	933.0
131.90	564.0
133.10	395.0
133.90	1452.0
135.00	3928.0
136.00	1721.0
137.00	2307.0
139.00	445.0
139.90	798.0
141.00	7244.0
142.00	2043.0
143.00	1337.0
144.10	422.0
145.00	374.0
145.90	1281.0
147.00	3712.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

148.00	10131.0
149.00	1751.0
150.00	476.0
150.90	796.0
151.90	429.0
153.10	2345.0
154.10	1580.0
155.00	3833.0
156.00	5245.0
157.10	1470.0
158.00	1096.0
159.00	1170.0
160.00	2433.0
161.00	2521.0
161.90	1051.0
165.00	3725.0
166.10	2109.0
167.00	14961.0
168.00	8479.0
169.00	951.0
169.90	533.0
171.00	817.0
171.90	1370.0
172.90	1545.0
174.00	3066.0
175.00	5584.0
176.00	665.0
176.90	2522.0
177.90	944.0
179.00	9942.0
180.00	6474.0
181.00	3065.0
182.30	426.0
184.10	838.0
185.00	3594.0
186.00	34976.0
187.00	9927.0
188.00	1249.0
189.00	2477.0
190.00	398.0
191.00	1414.0
192.00	3267.0
193.10	3383.0
194.00	855.0
194.80	486.0
196.00	5571.0
197.10	2173.0
198.00	216896.0
199.00	12268.0
199.90	1026.0
201.30	971.0
202.30	415.0
203.00	2036.0
204.00	9861.0
205.10	15304.0
206.00	62640.0
207.00	9496.0
207.90	2542.0
209.00	941.0
210.00	1103.0
210.90	2138.0
211.80	331.0
214.90	891.0
215.90	1535.0
217.00	18528.0
218.00	2259.0
221.00	9933.0
221.90	2655.0
223.00	3526.0
224.10	33944.0
225.00	9594.0
225.90	1148.0
227.00	18504.0
228.00	2205.0
228.90	3123.0
230.00	493.0
231.10	1240.0
233.90	1003.0
235.00	1406.0
236.00	773.0
237.00	1384.0
239.00	679.0
240.00	779.0
240.90	803.0
242.00	2084.0
243.00	2043.0
244.10	27416.0
245.10	3519.0
246.00	6822.0
246.90	1203.0
247.90	454.0
249.00	730.0
249.90	359.0
251.00	346.0
252.10	326.0
253.00	895.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

254.10	883.0
255.00	133120.0
256.00	19992.0
257.00	2397.0
258.00	9334.0
259.00	2235.0
261.00	374.0
265.00	3668.0
266.10	441.0
271.00	518.0
272.00	473.0
273.10	4343.0
274.00	10861.0
275.00	69688.0
276.00	9116.0
277.00	5764.0
278.00	725.0
281.00	522.0
283.00	859.0
283.90	353.0
285.00	957.0
288.90	557.0
292.00	359.0
293.00	1642.0
293.90	462.0
296.00	20784.0
297.00	3616.0
301.00	349.0
301.90	423.0
303.00	2777.0
304.00	476.0
307.90	329.0
313.90	914.0
315.00	2095.0
316.00	1504.0
321.00	772.0
323.00	6871.0
324.00	995.0
326.90	1009.0
328.00	797.0
332.00	563.0
332.90	499.0
334.00	4264.0
335.20	846.0
341.10	932.0
345.90	1415.0
346.90	429.0
351.90	2299.0
352.90	1771.0
354.00	2075.0
354.90	762.0
365.00	10746.0
366.00	1525.0
370.80	351.0
371.20	340.0
372.00	3843.0
373.00	608.0
383.00	961.0
384.00	388.0
391.10	390.0
402.00	1254.0
403.00	1322.0
403.90	703.0
421.00	1481.0
422.10	1521.0
423.10	10371.0
424.00	2567.0
441.00	36824.0
442.00	220672.0
443.00	41040.0
444.00	4022.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

m/z	Abundance
37.90	549.0
39.00	3738.0
39.90	761.0
41.00	637.0
43.00	526.0
44.00	2837.0
49.10	794.0
50.00	14904.0
51.10	54168.0
52.00	2609.0
55.10	385.0
56.10	1466.0
57.00	3343.0
61.10	740.0
62.00	638.0
63.00	2372.0
63.90	513.0
65.00	858.0
68.10	837.0
69.00	49800.0
70.00	335.0
73.00	1995.0
74.00	6292.0
75.00	9901.0
76.10	3119.0
77.10	57928.0
78.00	4435.0
79.00	4317.0
80.00	3311.0
81.00	4734.0
81.90	1109.0
83.10	1005.0
85.00	944.0
86.00	1039.0
86.90	477.0
91.10	1173.0
92.00	1072.0
93.00	8124.0
94.10	302.0
95.00	332.0
96.10	422.0
98.00	7467.0
99.00	5647.0
100.00	527.0
101.00	2461.0
103.00	566.0
104.00	1796.0
105.00	1770.0
105.80	498.0
107.00	18176.0
108.00	2273.0
109.10	379.0
110.00	28784.0
111.00	5149.0
112.00	508.0
116.00	1062.0
117.00	23120.0
117.90	1389.0
122.00	894.0
123.00	1266.0
123.90	1033.0
125.00	993.0
127.00	54832.0
128.00	4211.0
129.00	26208.0
130.00	2031.0
131.00	605.0
132.00	393.0
133.90	605.0
135.00	2055.0
136.00	761.0
137.00	1225.0
137.90	505.0
141.00	4328.0
142.00	1107.0
142.90	544.0
143.80	346.0
145.90	614.0
146.90	2752.0
148.00	4835.0
149.00	979.0
150.90	428.0
151.90	355.0
152.90	1296.0
153.90	1077.0
155.00	1940.0
156.10	2507.0
157.00	609.0
158.10	784.0
158.90	379.0
160.00	978.0
161.00	1861.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

161.90	687.0
163.10	311.0
165.00	1305.0
166.00	1059.0
167.00	8665.0
168.00	4007.0
169.00	666.0
170.10	348.0
170.80	383.0
172.10	861.0
173.00	731.0
174.00	1683.0
175.00	2829.0
175.90	672.0
177.00	1336.0
178.00	642.0
179.00	5407.0
180.00	3348.0
181.00	1493.0
184.10	522.0
185.00	2706.0
186.00	19280.0
187.00	6099.0
187.90	792.0
189.00	1605.0
190.00	306.0
191.00	726.0
191.90	1675.0
193.00	1850.0
193.90	511.0
196.00	3877.0
198.00	121744.0
198.90	7034.0
200.00	1480.0
201.40	565.0
203.00	1747.0
204.10	5314.0
205.00	8705.0
206.00	36008.0
207.10	5355.0
208.10	1716.0
209.00	745.0
210.20	565.0
210.90	1217.0
211.70	344.0
214.90	615.0
216.00	859.0
217.00	10364.0
218.00	1241.0
221.00	4990.0
221.70	1398.0
222.90	1731.0
224.00	19752.0
225.00	5350.0
226.00	595.0
227.00	10174.0
228.00	1927.0
229.00	1637.0
231.00	891.0
233.90	1030.0
235.00	633.0
235.90	612.0
237.10	915.0
239.80	370.0
241.00	675.0
241.90	1300.0
243.00	1130.0
244.00	16079.0
245.00	1753.0
246.00	3802.0
246.90	852.0
248.90	674.0
253.00	311.0
255.00	87304.0
256.00	12592.0
257.00	1129.0
258.00	6110.0
259.10	934.0
264.90	2662.0
265.80	438.0
271.90	329.0
273.00	2182.0
274.00	7277.0
275.00	37856.0
276.00	4968.0
277.00	3195.0
278.00	703.0
280.90	347.0
283.10	541.0
283.80	352.0
284.90	682.0
292.80	585.0
295.10	413.0
296.00	12916.0
297.00	1741.0
303.00	2301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32

304.00	359.0
314.00	431.0
314.90	1720.0
316.00	912.0
323.00	3535.0
323.90	1079.0
326.90	688.0
327.90	475.0
333.00	363.0
334.00	2880.0
335.00	680.0
340.90	542.0
345.90	914.0
346.90	306.0
351.90	1601.0
352.90	972.0
353.90	995.0
355.00	392.0
355.10	398.0
365.00	7344.0
365.90	959.0
371.00	525.0
372.00	2772.0
373.10	494.0
383.00	971.0
402.00	636.0
402.90	1432.0
404.00	328.0
420.90	1333.0
421.90	1144.0
423.00	8697.0
424.00	1748.0
441.00	24472.0
442.10	165568.0
443.10	32560.0
444.00	2277.0

Instrument :
BNA_N
ClientSampleId :
DFTPP32