

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122220\
 Data File : BN013243.D
 Acq On : 21 Dec 2020 18:40
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 22 01:05:39 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 16:11:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	130	0.00
2	1,4-Dioxane	0.400	0.440	-10.0	142	0.00
3	n-Nitrosodimethylamine	0.400	0.296	26.0#	88	0.00
4 S	2-Fluorophenol	0.400	0.375	6.3	105	0.00
5 S	Phenol-d6	0.400	0.409	-2.2	132	0.00
6	bis(2-Chloroethyl)ether	0.400	0.385	3.8	140	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	132	0.00
8 S	Nitrobenzene-d5	0.400	0.382	4.5	132	0.00
9	Naphthalene	0.400	0.357	10.8	117	-0.01
10	Hexachlorobutadiene	0.400	0.327	18.3	106	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.468	-17.0	156	0.00
12	2-Methylnaphthalene	0.400	0.308	23.0	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	121	0.00
14 S	2,4,6-Tribromophenol	0.400	0.376	6.0	126	0.00
15 S	2-Fluorobiphenyl	0.400	0.348	13.0	107	0.00
16	Acenaphthylene	0.400	0.361	9.8	111	0.00
17	Acenaphthene	0.400	0.358	10.5	110	0.00
18	Fluorene	0.400	0.371	7.3	116	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	135	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.377	5.8	124	-0.01
21	4-Bromophenyl-phenylether	0.400	0.167	58.3#	61	0.00
22	Hexachlorobenzene	0.400	0.327	18.3	106	0.00
23	Atrazine	0.400	0.335	16.3	123	-0.01
24	Pentachlorophenol	0.400	0.350	12.5	130	0.00
25	Phenanthrene	0.400	0.342	14.5	121	-0.01
26	Anthracene	0.400	0.343	14.2	125	0.00
27 SURR	Fluoranthene-d10	0.400	0.523	-30.8#	183	0.00
28	Fluoranthene	0.400	0.356	11.0	125	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	141	0.00
30	Pyrene	0.400	0.357	10.8	127	0.00
31 S	Terphenyl-d14	0.400	0.349	12.8	124	0.00
32	Benzo(a)anthracene	0.400	0.347	13.3	124	0.00
33	Chrysene	0.400	0.367	8.3	133	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.221	44.8#	118	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.356	11.0	116	0.00
36 I	Perylene-d12	0.400	0.400	0.0	135	0.00
37	Benzo(b)fluoranthene	0.400	0.357	10.8	122	0.00
38	Benzo(k)fluoranthene	0.400	0.336	16.0	121	0.00
39 C	Benzo(a)pyrene	0.400	0.363	9.3	122	0.00
40	Dibenzo(a,h)anthracene	0.400	0.344	14.0	118	0.00
41	Benzo(g,h,i)perylene	0.400	0.344	14.0	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4