

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122220\
 Data File : BN013260.D
 Acq On : 22 Dec 2020 05:30
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 22 06:04:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 18 16:11:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	0.00
2	1,4-Dioxane	0.400	0.427	-6.7	132	0.00
3	n-Nitrosodimethylamine	0.400	0.283	29.3#	80	0.00
4 S	2-Fluorophenol	0.400	0.320	20.0	86	0.00
5 S	Phenol-d6	0.400	0.334	16.5	104	0.00
6	bis(2-Chloroethyl)ether	0.400	0.299	25.3#	104	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.372	7.0	104	0.00
9	Naphthalene	0.400	0.322	19.5	86	0.00
10	Hexachlorobutadiene	0.400	0.319	20.3	84	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.481	-20.2	130	0.00
12	2-Methylnaphthalene	0.400	0.314	21.5	86	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.400	0.319	20.3	102	0.00
15 S	2-Fluorobiphenyl	0.400	0.302	24.5	89	0.01
16	Acenaphthylene	0.400	0.302	24.5	89	0.00
17	Acenaphthene	0.400	0.316	21.0	93	0.01
18	Fluorene	0.400	0.316	21.0	95	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	124	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.345	13.8	94	0.00
21	4-Bromophenyl-phenylether	0.400	0.190	52.5#	64	0.00
22	Hexachlorobenzene	0.400	0.305	23.8	91	0.00
23	Atrazine	0.400	0.279	30.3#	94	0.00
24	Pentachlorophenol	0.400	0.288	28.0#	98	0.01
25	Phenanthrene	0.400	0.304	24.0	98	0.00
26	Anthracene	0.400	0.291	27.3#	98	0.00
27 SURR	Fluoranthene-d10	0.400	0.456	-14.0	147	0.00
28	Fluoranthene	0.400	0.310	22.5	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	125	0.00
30	Pyrene	0.400	0.315	21.3	99	0.00
31 S	Terphenyl-d14	0.400	0.315	21.3	99	0.00
32	Benzo(a)anthracene	0.400	0.293	26.8#	93	0.00
33	Chrysene	0.400	0.335	16.3	107	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.086	78.5#	75	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.335	16.3	96	0.00
36 I	Perylene-d12	0.400	0.400	0.0	121	0.00
37	Benzo(b)fluoranthene	0.400	0.325	18.8	99	0.00
38	Benzo(k)fluoranthene	0.400	0.315	21.3	101	0.00
39 C	Benzo(a)pyrene	0.400	0.326	18.5	98	0.00
40	Dibenzo(a,h)anthracene	0.400	0.319	20.3	97	0.00
41	Benzo(g,h,i)perylene	0.400	0.328	18.0	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4