

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122220\
 Data File : BN013276.D
 Acq On : 22 Dec 2020 17:00
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 22 17:31:38 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN121620.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 22 10:46:43 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.538	-34.5#	165	0.00
3	n-Nitrosodimethylamine	0.400	0.299	25.3#	84	0.00
4 S	2-Fluorophenol	0.400	0.365	8.8	97	0.00
5 S	Phenol-d6	0.400	0.330	17.5	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.291	27.3#	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	125	0.00
8 S	Nitrobenzene-d5	0.400	0.303	24.3	99	0.01
9	Naphthalene	0.400	0.316	21.0	98	0.00
10	Hexachlorobutadiene	0.400	0.320	20.0	98	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.479	-19.7	150	0.00
12	2-Methylnaphthalene	0.400	0.304	24.0	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	134	0.00
14 S	2,4,6-Tribromophenol	0.400	0.325	18.8	121	0.00
15 S	2-Fluorobiphenyl	0.400	0.302	24.5	103	0.00
16	Acenaphthylene	0.400	0.306	23.5	104	0.00
17	Acenaphthene	0.400	0.313	21.8	107	0.00
18	Fluorene	0.400	0.313	21.8	109	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	141	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.353	11.8	112	0.01
21	4-Bromophenyl-phenylether	0.400	0.175	56.3#	67	0.00
22	Hexachlorobenzene	0.400	0.299	25.3#	101	0.00
23	Atrazine	0.400	0.282	29.5#	108	0.00
24	Pentachlorophenol	0.400	0.299	25.3#	115	0.01
25	Phenanthrene	0.400	0.304	24.0	112	-0.01
26	Anthracene	0.400	0.286	28.5#	109	0.01
27 SURR	Fluoranthene-d10	0.400	0.445	-11.2	162	0.00
28	Fluoranthene	0.400	0.302	24.5	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	137	0.00
30	Pyrene	0.400	0.316	21.0	109	0.00
31 S	Terphenyl-d14	0.400	0.314	21.5	108	0.00
32	Benzo(a)anthracene	0.400	0.281	29.8#	97	0.00
33	Chrysene	0.400	0.337	15.8	119	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.072	82.0#	79	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.322	19.5	102	0.00
36 I	Perylene-d12	0.400	0.400	0.0	127	0.00
37	Benzo(b)fluoranthene	0.400	0.341	14.8	109	0.00
38	Benzo(k)fluoranthene	0.400	0.305	23.8	103	0.00
39 C	Benzo(a)pyrene	0.400	0.334	16.5	105	0.00
40	Dibenzo(a,h)anthracene	0.400	0.321	19.8	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.328	18.0	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4