


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122223\
 Data File : BN029182.D
 Acq On : 22 Dec 2023 18:06
 Operator : MA/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Dec 23 03:23:24 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 23 03:20:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	948	0.400	ng	0.00
7) Naphthalene-d8	10.562	136	1725	0.400	ng	-0.01
13) Acenaphthene-d10	14.414	164	927	0.400	ng	-0.01
19) Phenanthrene-d10	17.180	188	1511	0.400	ng	0.00
29) Chrysene-d12	21.384	240	640	0.400	ng	0.00
35) Perylene-d12	23.712	264	786	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	6895	3.038	ng	0.00
5) Phenol-d6	6.980	99	7955	3.190	ng	0.00
8) Nitrobenzene-d5	8.950	82	7310	3.252	ng	-0.01
11) 2-Methylnaphthalene-d10	12.158	152	9662	3.697	ng	0.00
14) 2,4,6-Tribromophenol	15.914	330	2163	2.669	ng	-0.01
15) 2-Fluorobiphenyl	13.046	172	17501	3.233	ng	-0.01
27) Fluoranthene-d10	19.215	212	10170	2.781	ng	0.00
31) Terphenyl-d14	19.828	244	5851	3.328	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.318	88	2725	3.051	ng	92
3) n-Nitrosodimethylamine	3.658	42	2235	3.647	ng	# 74
6) bis(2-Chloroethyl)ether	7.226	93	5907	3.048	ng	92
9) Naphthalene	10.615	128	17274	3.073	ng	# 89
10) Hexachlorobutadiene	10.882	225	5216	2.820	ng	# 99
12) 2-Methylnaphthalene	12.234	142	12665	3.447	ng	96
16) Acenaphthylene	14.136	152	20401	3.123	ng	99
17) Acenaphthene	14.479	154	12263	3.129	ng	98
18) Fluorene	15.473	166	17463	3.305	ng	98
20) 4,6-Dinitro-2-methylph...	15.567	198	2105	3.473	ng	# 40
21) 4-Bromophenyl-phenylether	16.374	248	4746	3.037	ng	# 79
22) Hexachlorobenzene	16.461	284	5442	2.991	ng	98
23) Atrazine	16.659	200	4129	3.150	ng	# 87
24) Pentachlorophenol	16.833	266	3386	3.139	ng	96
25) Phenanthrene	17.218	178	21345	3.424	ng	99
26) Anthracene	17.305	178	20973	3.480	ng	100
28) Fluoranthene	19.247	202	18927	2.936	ng	# 99
30) Pyrene	19.610	202	18685	3.706	ng	99
32) Benzo(a)anthracene	21.366	228	12473	3.544	ng	93
33) Chrysene	21.420	228	11596	3.332	ng	92
34) Bis(2-ethylhexyl)phtha...	21.312	149	13390	3.212	ng	98
36) Indeno(1,2,3-cd)pyrene	26.086	276	18390	3.633	ng	# 86
37) Benzo(b)fluoranthene	23.004	252	14048	3.514	ng	# 75
38) Benzo(k)fluoranthene	23.054	252	14232	3.636	ng	# 75
39) Benzo(a)pyrene	23.607	252	12803	3.589	ng	# 66
40) Dibenzo(a,h)anthracene	26.115	278	14306	3.674	ng	# 70
41) Benzo(g,h,i)perylene	26.817	276	15354	3.581	ng	# 85





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.775 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 948
Ion Ratio Lower Upper
152 100
150 149.2 117.5 176.3
115 57.4 46.3 69.5



#2
1,4-Dioxane
Concen: 3.051 ng
RT: 3.318 min Scan# 33
Delta R.T. -0.008 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion: 88 Resp: 2725
Ion Ratio Lower Upper
88 100
43 22.1 19.1 28.7
58 41.5 28.2 42.4









#3
n-Nitrosodimethylamine
Concen: 3.647 ng
RT: 3.658 min Scan# 80
Delta R.T. -0.029 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 2235
Ion Ratio Lower Upper
42 100
74 148.1 95.1 142.7#
44 2.9 7.6 11.4#



#4
2-Fluorophenol
Concen: 3.038 ng
RT: 5.384 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion: 112 Resp: 6895
Ion Ratio Lower Upper
112 100
64 49.0 37.8 56.6
63 37.8 29.4 44.0









#5
Phenol-d6
Concen: 3.190 ng
RT: 6.980 min Scan# 54
Delta R.T. -0.008 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 7955
Ion Ratio Lower Upper
99 100
42 18.1 16.8 25.2
71 58.6 43.4 65.2



#6
bis(2-Chloroethyl)ether
Concen: 3.048 ng
RT: 7.226 min Scan# 574
Delta R.T. -0.015 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion: 93 Resp: 5907
Ion Ratio Lower Upper
93 100
63 65.9 58.4 87.6
95 33.2 29.2 43.8





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.562 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2

Tgt Ion:136 Resp: 1725





Ion Ratio Lower Upper

136 100

137 15.1 12.1 18.1

54 10.7 7.2 10.8

68 8.1 6.0 9.0



#8

Nitrobenzene-d5

Concen: 3.252 ng

RT: 8.950 min Scan# 784

Delta R.T. -0.011 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion: 82 Resp: 7310


Ion Ratio Lower Upper

82 100

128 35.2 34.5 51.7

54 32.8 33.4 50.0#









#9
Naphthalene
Concen: 3.073 ng
RT: 10.615 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 17274
Ion Ratio Lower Upper
128 100
129 11.1 13.0 19.6#
127 13.8 14.4 21.6#



#10
Hexachlorobutadiene
Concen: 2.820 ng
RT: 10.882 min Scan# 965
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:225 Resp: 5216
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.4 75.6









#11
2-Methylnaphthalene-d10
Concen: 3.697 ng
RT: 12.158 min Scan# 1195
Delta R.T. -0.008 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 9662
Ion Ratio Lower Upper
152 100
151 21.2 17.4 26.0



#12
2-Methylnaphthalene
Concen: 3.447 ng
RT: 12.234 min Scan# 1215
Delta R.T. -0.008 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:142 Resp: 12665
Ion Ratio Lower Upper
142 100
141 88.4 71.9 107.9
115 51.7 46.1 69.1









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.414 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 927
Ion Ratio Lower Upper
164 100
162 114.0 88.2 132.2
160 58.5 44.9 67.3



#14
2,4,6-Tribromophenol
Concen: 2.669 ng
RT: 15.914 min Scan# 1669
Delta R.T. -0.013 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:330 Resp: 2163
Ion Ratio Lower Upper
330 100
332 94.7 77.6 116.4
141 58.9 40.2 60.2









#15
2-Fluorobiphenyl
Concen: 3.233 ng
RT: 13.046 min Scan# 1407
Delta R.T. -0.011 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 17501
Ion Ratio Lower Upper
172 100
171 36.0 31.4 47.0
170 24.7 22.8 34.2



#16
Acenaphthylene
Concen: 3.123 ng
RT: 14.136 min Scan# 1508
Delta R.T. -0.011 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06


Tgt Ion:152 Resp: 20401
Ion Ratio Lower Upper
152 100
151 19.4 15.1 22.7
153 12.9 10.3 15.5








#17
Acenaphthene
Concen: 3.129 ng
RT: 14.479 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:154 Resp: 12263
Ion Ratio Lower Upper
154 100
153 119.1 96.8 145.2
152 56.7 46.1 69.1



#18
Fluorene
Concen: 3.305 ng
RT: 15.473 min Scan# 1633
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:166 Resp: 17463
Ion Ratio Lower Upper
166 100
165 102.3 80.0 120.0
167 12.8 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.180 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 1511

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.4 8.9 13.3



#20

4,6-Dinitro-2-methylphenol

Concen: 3.473 ng

RT: 15.567 min Scan# 1641

Delta R.T. -0.013 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:198 Resp: 2105


Ion Ratio Lower Upper

198 100

51 17.6 56.0 84.0#

105 28.9 61.6 92.4#









#21
4-Bromophenyl-phenylether
Concen: 3.037 ng
RT: 16.374 min Scan# 1706
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 4746
Ion Ratio Lower Upper
248 100
250 95.1 75.7 113.5
141 71.1 89.4 134.0#



#22
Hexachlorobenzene
Concen: 2.991 ng
RT: 16.461 min Scan# 1713
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:284 Resp: 5442
Ion Ratio Lower Upper
284 100
142 59.0 47.4 71.0
249 33.4 28.8 43.2





#23

Atrazine

Concen: 3.150 ng

RT: 16.659 min Scan# 1729

Delta R.T. -0.013 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:200 Resp: 4129

Ion Ratio Lower Upper

200 100

173 31.6 30.5 45.7

215 37.6 37.8 56.6#



#24

Pentachlorophenol

Concen: 3.139 ng

RT: 16.833 min Scan# 1743

Delta R.T. -0.000 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:266 Resp: 3386


Ion Ratio Lower Upper

266 100

264 61.8 52.8 79.2

268 62.7 51.4 77.0





#25

Phenanthrene

Concen: 3.424 ng

RT: 17.218 min Scan# 1774

Delta R.T. -0.000 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:178 Resp: 21345

Ion Ratio Lower Upper

178 100

176 20.7 16.6 25.0

179 14.6 12.2 18.4



#26

Anthracene

Concen: 3.480 ng

RT: 17.305 min Scan# 1781

Delta R.T. -0.013 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:178 Resp: 20973


Ion Ratio Lower Upper

178 100

176 19.9 15.8 23.8

179 14.9 11.9 17.9









#27
Fluoranthene-d10
Concen: 2.781 ng
RT: 19.215 min Scan# 2030
Delta R.T. -0.005 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 10170
Ion Ratio Lower Upper
212 100
106 11.4 9.6 14.4
104 7.8 5.8 8.6



#28
Fluoranthene
Concen: 2.936 ng
RT: 19.247 min Scan# 2037
Delta R.T. -0.005 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:202 Resp: 18927
Ion Ratio Lower Upper
202 100
101 6.7 4.5 6.7#
203 16.8 13.7 20.5





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 2458

Delta R.T. -0.000 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 640

Ion Ratio Lower Upper

240 100

120 27.9 23.9 35.9

236 42.4 31.3 46.9



#30

Pyrene

Concen: 3.706 ng

RT: 19.610 min Scan# 2115

Delta R.T. -0.005 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:202 Resp: 18685


Ion Ratio Lower Upper

202 100

200 22.8 17.4 26.2

203 17.4 13.8 20.8





#31

Terphenyl-d14

Concen: 3.328 ng

RT: 19.828 min Scan# 2162

Delta R.T. -0.005 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:244 Resp: 5851

Ion Ratio Lower Upper

244 100

212 17.4 18.9 28.3#

122 21.0 22.5 33.7#



#32

Benzo(a)anthracene

Concen: 3.544 ng

RT: 21.366 min Scan# 2456

Delta R.T. -0.009 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:228 Resp: 12473


Ion Ratio Lower Upper

228 100

226 30.2 27.8 41.6

229 19.6 18.0 27.0









#33
Chrysene
Concen: 3.332 ng
RT: 21.420 min Scan# 2462
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 11596
Ion Ratio Lower Upper
228 100
226 34.3 32.0 48.0
229 18.9 17.7 26.5



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.212 ng
RT: 21.312 min Scan# 2450
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:149 Resp: 13390
Ion Ratio Lower Upper
149 100
167 25.0 20.8 31.2
279 2.1 1.7 2.5





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.712 min Scan# 3017

Delta R.T. 0.002 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 786

Ion Ratio Lower Upper

264 100

260 36.8 27.1 40.7

265 125.3 99.4 149.2



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.633 ng

RT: 26.086 min Scan# 3831

Delta R.T. -0.009 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:276 Resp: 18390


Ion Ratio Lower Upper

276 100

138 24.1 6.9 10.3#

277 24.5 18.4 27.6





#37

Benzo(b)fluoranthene

Concen: 3.514 ng

RT: 23.004 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 14048

Ion Ratio Lower Upper

252 100

253 24.1 33.4 50.0#

125 9.6 14.6 21.8#



#38

Benzo(k)fluoranthene

Concen: 3.636 ng

RT: 23.054 min Scan# 2794

Delta R.T. 0.002 min

Lab File: BN029182.D

Acq: 22 Dec 2023 18:06




Tgt Ion:252 Resp: 14232


Ion Ratio Lower Upper

252 100

253 23.9 33.3 49.9#


125 9.3 14.0 21.0#








#39
Benzo(a)pyrene
Concen: 3.589 ng
RT: 23.607 min Scan# 2981
Delta R.T. -0.003 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:252 Resp: 12803
Ion Ratio Lower Upper
252 100
253 23.7 39.7 59.5#
125 10.5 17.9 26.9#



#40
Dibenzo(a,h)anthracene
Concen: 3.674 ng
RT: 26.115 min Scan# 3841
Delta R.T. -0.000 min
Lab File: BN029182.D
Acq: 22 Dec 2023 18:06

Tgt Ion:278 Resp: 14306
Ion Ratio Lower Upper
278 100
139 17.3 21.8 32.6#
279 25.3 39.9 59.9#





#41

Benzo(g,h,i)perylene

Concen: 3.581 ng

RT: 26.817 min Scan# 40

Delta R.T. 0.002 min

Lab File: BN029182.D


Acq: 22 Dec 2023 18:06

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2



Tgt Ion:276 Resp: 15354

Ion Ratio Lower Upper

276 100

277 23.9 24.2 36.4#

138 21.5 25.5 38.3#

