

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122322\
 Data File : BN023372.D
 Acq On : 23 Dec 2022 14:07
 Operator : CG/JU
 Sample : N6164-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-04-147-166-122022

Quant Time: Dec 23 16:39:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 16:36:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.992	152	6554	0.400	ng	0.00
7) Naphthalene-d8	10.808	136	21910	0.400	ng	0.00
13) Acenaphthene-d10	14.645	164	13760	0.400	ng	0.00
19) Phenanthrene-d10	17.390	188	30295	0.400	ng	# 0.00
29) Chrysene-d12	21.580	240	22301	0.400	ng	0.00
35) Perylene-d12	24.033	264	15547	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.536	112	1701	0.139	ng	0.00
5) Phenol-d6	7.154	99	1624	0.105	ng	0.00
8) Nitrobenzene-d5	9.153	82	4657	0.323	ng	0.00
11) 2-Methylnaphthalene-d10	12.397	152	9314	0.251	ng	0.00
14) 2,4,6-Tribromophenol	16.136	330	2002	0.401	ng	0.00
15) 2-Fluorobiphenyl	13.266	172	15507	0.282	ng	0.00
27) Fluoranthene-d10	19.422	212	26024	0.367	ng	0.00
31) Terphenyl-d14	20.013	244	17220	0.476	ng	0.00
Target Compounds						
18) Fluorene	15.693	166	1371	0.030	ng	# 95
25) Phenanthrene	17.427	178	59216	0.656	ng	100
28) Fluoranthene	19.452	202	84324	0.872	ng	100
30) Pyrene	19.814	202	38507	0.472	ng	97
32) Benzo(a)anthracene	21.562	228	1702	0.024	ng	# 77
33) Chrysene	21.616	228	2349	0.029	ng	# 75
34) Bis(2-ethylhexyl)phtha...	21.482	149	4076	0.134	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122322\
Data File : BN023372.D
Acq On : 23 Dec 2022 14:07
Operator : CG/JU
Sample : N6164-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-04-147-166-122022

Quant Time: Dec 23 16:39:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 23 16:36:53 2022
Response via : Initial Calibration

