

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122322\
 Data File : BN023373.D
 Acq On : 23 Dec 2022 15:58
 Operator : CG/JU
 Sample : N6164-06
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-04-165-186-122022

Quant Time: Dec 23 16:44:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 23 16:36:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.991	152	6856	0.400	ng	0.00
7) Naphthalene-d8	10.808	136	22889	0.400	ng	0.00
13) Acenaphthene-d10	14.645	164	14508	0.400	ng	0.00
19) Phenanthrene-d10	17.389	188	32049	0.400	ng	0.00
29) Chrysene-d12	21.580	240	22248	0.400	ng	# 0.00
35) Perylene-d12	24.036	264	15621	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.536	112	1573	0.123	ng	0.00
5) Phenol-d6	7.154	99	1395	0.086	ng	0.00
8) Nitrobenzene-d5	9.153	82	4128	0.274	ng	0.00
11) 2-Methylnaphthalene-d10	12.397	152	8693	0.224	ng	0.00
14) 2,4,6-Tribromophenol	16.136	330	2247	0.427	ng	0.00
15) 2-Fluorobiphenyl	13.266	172	14108	0.243	ng	0.00
27) Fluoranthene-d10	19.418	212	23675	0.316	ng	0.00
31) Terphenyl-d14	20.013	244	14274	0.395	ng	0.00
Target Compounds						
25) Phenanthrene	17.427	178	16662	0.174	ng	Qvalue 100
28) Fluoranthene	19.452	202	15132	0.148	ng	99
30) Pyrene	19.814	202	8488	0.104	ng	96
34) Bis(2-ethylhexyl)phtha...	21.482	149	8471	0.280	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122322\
Data File : BN023373.D
Acq On : 23 Dec 2022 15:58
Operator : CG/JU
Sample : N6164-06
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-04-165-186-122022

Quant Time: Dec 23 16:44:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN120822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 23 16:36:53 2022
Response via : Initial Calibration

