

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122324\
 Data File : BN035795.D
 Acq On : 23 Dec 2024 22:52
 Operator : RC/JU
 Sample : PB165667BL
 Misc :
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB165667BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/24/2024
 Supervised By :mohammad ahmed 12/26/2024

Quant Time: Dec 24 00:13:30 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 23:06:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.264	152	4926	0.400	ng	0.00
7) Naphthalene-d8	10.009	136	11442	0.400	ng	-0.01
13) Acenaphthene-d10	13.924	164	7253	0.400	ng	-0.01
19) Phenanthrene-d10	16.711	188	12605	0.400	ng	# 0.01
29) Chrysene-d12	20.956	240	9944	0.400	ng	# 0.00
35) Perylene-d12	23.038	264	9785	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.932	112	3783	0.307	ng	0.00
5) Phenol-d6	6.484	99	3587	0.242	ng	0.00
8) Nitrobenzene-d5	8.408	82	2999m	0.429	ng	0.00
11) 2-Methylnaphthalene-d10	11.626	152	6566	0.367	ng	0.00
14) 2,4,6-Tribromophenol	15.475	330	426	0.083	ng	0.02
15) 2-Fluorobiphenyl	12.549	172	9720	0.355	ng	0.00
27) Fluoranthene-d10	18.761	212	13649	0.382	ng	0.00
31) Terphenyl-d14	19.388	244	11020	0.562	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122324\
Data File : BN035795.D
Acq On : 23 Dec 2024 22:52
Operator : RC/JU
Sample : PB165667BL
Misc :
ALS Vial : 77 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB165667BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/24/2024
Supervised By :mohammad ahmed 12/26/2024

Quant Time: Dec 24 00:13:30 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 23:06:19 2024
Response via : Initial Calibration

