

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122324\
 Data File : BN035836.D
 Acq On : 25 Dec 2024 01:05
 Operator : RC/JU
 Sample : P5376-19
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TT188S1-20241219

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/26/2024
 Supervised By :mohammad ahmed 12/26/2024

Quant Time: Dec 25 02:35:26 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 23:06:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.257	152	3201	0.400	ng	-0.01
7) Naphthalene-d8	10.009	136	7225	0.400	ng	-0.01
13) Acenaphthene-d10	13.925	164	4592	0.400	ng	-0.01
19) Phenanthrene-d10	16.698	188	9284	0.400	ng	# 0.00
29) Chrysene-d12	20.947	240	7823	0.400	ng	# 0.00
35) Perylene-d12	23.032	264	7346	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	4.925	112	813	0.101	ng	-0.01
5) Phenol-d6	6.485	99	377	0.039	ng	0.00
8) Nitrobenzene-d5	8.408	82	1685m	0.382	ng	0.00
11) 2-Methylnaphthalene-d10	11.626	152	3842	0.340	ng	0.00
14) 2,4,6-Tribromophenol	15.453	330	593	0.182	ng	0.00
15) 2-Fluorobiphenyl	12.544	172	6127	0.353	ng	0.00
27) Fluoranthene-d10	18.752	212	11461	0.435	ng	0.00
31) Terphenyl-d14	19.379	244	8539	0.553	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.982	88	618	0.202	ng	# 77
34) Bis(2-ethylhexyl)phtha...	20.893	149	434	0.040	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122324\
Data File : BN035836.D
Acq On : 25 Dec 2024 01:05
Operator : RC/JU
Sample : P5376-19
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TT188S1-20241219

Quant Time: Dec 25 02:35:26 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 23:06:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/26/2024
Supervised By :mohammad ahmed 12/26/2024

