

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122618\
 Data File : BN004216.D
 Acq On : 26 Dec 2018 22:59
 Operator : JU/SJ
 Sample : J6489-14
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 F9F77

Quant Time: Dec 27 05:37:56 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 27 04:46:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.82	152	25396	20.00	ng/ul	0.00
18) Naphthalene-d8	10.62	136	126158	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.46	164	81663	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.21	188	197459	20.00	ng/ul	0.00
77) Chrysene-d12	21.40	240	245061	20.00	ng/ul	0.00
85) Perylene-d12	23.72	264	278419	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	0.00	96	0	0.00	ng/uL	
5) Phenol-d5	6.99	99	10687	4.28	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.16	67	29787	21.92	ng/ul	0.00
9) 2-Chlorophenol-d4	7.36	132	35005	17.89	ng/ul	0.00
13) 4-Methylphenol-d8	8.53	113	22664	10.83	ng/ul	0.00
19) Nitrobenzene-d5	8.99	128	23580	23.59	ng/ul	0.00
22) 2-Nitrophenol-d4	9.71	143	27916	24.66	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.25	165	45243	21.79	ng/ul	0.00
29) 4-Chloroaniline-d4	0.00	131	0	0.00	ng/ul	
43) Dimethylphthalate-d6	13.87	166	203462	27.91	ng/ul	0.00
46) Acenaphthylene-d8	14.16	160	219969	25.43	ng/ul	0.00
51) 4-Nitrophenol-d4	14.70	143	6227	4.13	ng/ul	0.02
57) Fluorene-d10	15.46	176	171970	27.28	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.59	200	34897	24.54	ng/ul	0.00
70) Anthracene-d10	17.31	188	298253	29.89	ng/ul	0.00
78) Pyrene-d10	19.60	212	341732	31.17	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.57	264	453577	30.03	ng/ul	0.00


Target Compounds					Ovalue	
33) 4-Chloro-3-methylphenol	11.91	107	6114	2.630	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN122618\
Data File : BN004216.D
Acq On : 26 Dec 2018 22:59
Operator : JU/SJ
Sample : J6489-14
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
F9F77

Quant Time: Dec 27 05:37:56 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 27 04:46:35 2018
Response via : Initial Calibration





#33

4-Chloro-3-methylphenol

Concen: 2.630 ng/ul

RT: 11.91 min Scan# 1517

Delta R.T. 0.00 min

Lab File: BN004216.D

Acq: 26 Dec 2018 22:59

Instrument :

BNA_N

ClientSampleId :

F9F77



Tot Ion:107 Resp: 6114

Ion Ratio Lower Upper

107 100

144 25.7 21.1 31.7

142 82.7 65.0 97.4



Abundance

Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E

