

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122624\
 Data File : BN035850.D
 Acq On : 26 Dec 2024 13:07
 Operator : RC/JU
 Sample : P5374-03
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-190A-GW-718-720

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/27/2024
 Supervised By :mohammad ahmed 12/30/2024

Quant Time: Dec 27 00:40:58 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 19 23:06:19 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.257	152	3717	0.400	ng	-0.01
7) Naphthalene-d8	9.999	136	9841	0.400	ng	#-0.02
13) Acenaphthene-d10	13.914	164	7032	0.400	ng	-0.02
19) Phenanthrene-d10	16.686	188	14733	0.400	ng	#-0.01
29) Chrysene-d12	20.938	240	12545	0.400	ng	#-0.02
35) Perylene-d12	23.018	264	12258	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	4.924	112	452	0.049	ng	-0.01
5) Phenol-d6	6.470	99	406	0.036	ng	-0.02
8) Nitrobenzene-d5	8.397	82	1453m	0.242	ng	-0.01
11) 2-Methylnaphthalene-d10	11.600	152	4322	0.281	ng	-0.02
14) 2,4,6-Tribromophenol	15.443	330	397	0.080	ng	-0.01
15) 2-Fluorobiphenyl	12.518	172	5672	0.213	ng	-0.03
27) Fluoranthene-d10	18.743	212	9176	0.220	ng	-0.02
31) Terphenyl-d14	19.375	244	7214	0.292	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	2.974	88	1795	0.505	ng	# 95
34) Bis(2-ethylhexyl)phtha...	20.893	149	19704	1.137	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122624\
Data File : BN035850.D
Acq On : 26 Dec 2024 13:07
Operator : RC/JU
Sample : P5374-03
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-190A-GW-718-720

Quant Time: Dec 27 00:40:58 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN112724.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 19 23:06:19 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/27/2024
Supervised By :mohammad ahmed 12/30/2024

