


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122718\
 Data File : BN004253.D
 Acq On : 28 Dec 2018 08:03
 Operator : JU/SJ
 Sample : DFTPP77
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP77

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN121218MA.M
 Title : SVOA CALIBRATION
 Last Update : Fri Dec 28 03:12:04 2018



AutoFind: Scans 2486, 2487, 2488; Background Corrected with Scan 2480

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	26.9	33490	PASS
68	69	0.00	2	0.8	293	PASS
69	198	0.00	100	30.8	38325	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	42.3	52704	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	124520	PASS
199	198	5	9	6.7	8357	PASS
275	198	10	60	31.9	39720	PASS
365	198	1	100	3.9	4798	PASS
441	198	0.01	100	15.5	19328	PASS
442	198	50	100	99.8	124221	PASS
443	442	15	24	19.2	23813	PASS

m/z	Abundance
44.00	382.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
38.00	466.0
39.00	2872.0
49.00	315.0
50.00	9406.0
51.00	37496.0
52.00	1925.0
56.00	1130.0
57.00	2732.0
61.00	515.0
61.90	492.0
63.00	1603.0
65.00	805.0
69.00	42720.0
72.90	359.0
74.00	4152.0
75.00	6653.0
76.00	2304.0
77.00	48104.0
78.00	3466.0
79.00	3224.0
80.00	2457.0
81.00	3439.0
82.00	747.0
83.00	775.0
85.00	664.0
86.00	974.0
87.00	413.0
91.00	746.0
92.00	886.0
93.00	5242.0
94.00	376.0
98.00	4050.0
99.00	3261.0
101.00	1939.0
103.00	667.0
104.00	1246.0
104.90	1155.0
106.10	415.0
107.00	14468.0
108.00	2228.0
109.10	534.0
110.00	29168.0
111.00	4334.0
112.00	455.0
116.00	819.0
117.00	10688.0
118.00	898.0
122.00	984.0
123.00	1513.0
123.90	663.0
125.00	679.0
127.00	56184.0
128.00	4388.0
129.00	20776.0
130.00	1751.0
131.00	339.0
134.00	613.0
135.00	1697.0
136.00	710.0
137.00	839.0
141.00	2701.0
142.00	905.0
142.90	559.0
145.90	478.0
147.00	1312.0
148.00	2890.0
148.90	650.0
151.10	317.0
153.00	851.0
154.00	698.0
155.00	1557.0
156.00	2244.0
157.10	458.0
157.90	485.0
159.00	404.0
160.00	838.0
161.00	1232.0
161.90	418.0
165.00	1044.0
166.00	830.0
167.00	5076.0
168.00	2627.0
169.00	487.0
171.90	481.0
172.90	644.0
174.00	1134.0
175.00	2231.0
176.00	708.0
177.00	1094.0
178.00	354.0
179.00	4122.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

180.00	2858.0
181.00	1321.0
185.00	2094.0
186.00	16118.0
187.00	4544.0
188.00	537.0
189.00	1008.0
191.00	500.0
192.00	1241.0
193.00	1390.0
196.00	3912.0
198.00	123840.0
199.00	8533.0
200.00	591.0
201.40	662.0
202.90	714.0
204.00	3962.0
205.00	6943.0
206.00	30664.0
207.00	4057.0
208.00	1028.0
210.30	603.0
210.90	1169.0
214.90	345.0
216.00	585.0
217.00	7510.0
218.00	998.0
221.00	5952.0
223.00	1764.0
224.00	16720.0
225.00	4434.0
226.00	534.0
227.00	6287.0
228.00	901.0
229.00	1479.0
231.00	697.0
234.00	417.0
235.00	488.0
235.90	313.0
237.00	599.0
241.00	376.0
242.00	833.0
243.00	939.0
244.00	14006.0
245.00	1918.0
246.00	2474.0
247.00	492.0
249.00	524.0
255.00	72944.0
256.00	10909.0
257.00	903.0
258.00	3734.0
258.90	645.0
265.00	1528.0
273.00	2391.0
274.00	5806.0
275.00	36656.0
276.00	5003.0
277.00	2594.0
278.00	401.0
283.00	324.0
285.00	503.0
293.00	619.0
296.00	8610.0
297.00	1223.0
303.00	1091.0
304.00	312.0
314.00	459.0
315.00	1016.0
316.00	678.0
323.00	3251.0
324.00	609.0
327.00	549.0
327.90	345.0
334.00	2035.0
335.00	560.0
341.00	408.0
346.00	780.0
352.00	974.0
353.00	674.0
354.10	1111.0
365.00	4042.0
366.00	579.0
372.10	1745.0
373.00	428.0
382.90	530.0
402.00	721.0
403.00	944.0
404.00	396.0
421.00	786.0
422.00	883.0
423.10	5976.0
424.00	1450.0
441.10	14537.0
442.10	94216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

443.1018120.0

444.101853.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
38.00	469.0
39.00	2957.0
44.00	307.0
49.00	302.0
50.00	9955.0
51.00	39320.0
52.00	1999.0
56.00	1196.0
57.00	2919.0
61.00	552.0
62.00	619.0
63.00	1732.0
65.00	889.0
68.10	880.0
69.00	45200.0
73.00	361.0
74.00	4418.0
75.00	7061.0
76.10	2432.0
77.00	51648.0
78.00	3482.0
79.00	3407.0
80.00	2536.0
81.00	3579.0
82.00	895.0
83.00	813.0
85.00	740.0
86.00	1030.0
87.00	487.0
91.00	860.0
92.00	954.0
93.00	5524.0
94.00	392.0
98.00	4290.0
99.00	3447.0
101.00	2058.0
103.00	772.0
104.00	1375.0
105.00	1208.0
107.00	15536.0
108.00	2409.0
110.00	31808.0
111.00	4774.0
112.00	555.0
116.00	911.0
117.00	11601.0
118.00	904.0
122.00	1123.0
123.00	1702.0
124.00	805.0
125.00	773.0
127.00	62432.0
128.00	4884.0
129.00	22984.0
130.00	2065.0
131.00	383.0
134.00	730.0
135.00	1847.0
136.00	784.0
137.00	979.0
141.00	2943.0
142.00	1132.0
143.00	655.0
146.00	619.0
147.00	1602.0
148.00	3300.0
149.00	749.0
151.20	454.0
153.00	981.0
154.00	761.0
155.00	1707.0
156.00	2652.0
157.10	520.0
158.00	634.0
158.90	435.0
160.00	932.0
161.00	1475.0
162.00	455.0
165.00	1069.0
166.00	972.0
167.00	6138.0
168.00	3005.0
169.00	528.0
172.00	538.0
173.00	758.0
174.00	1322.0
175.00	2542.0
176.00	864.0
177.00	1188.0
177.90	414.0
179.00	4816.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

180.00	3277.0
181.00	1582.0
184.00	388.0
185.00	2326.0
186.00	18688.0
187.00	5386.0
188.00	492.0
189.00	1109.0
191.00	546.0
192.00	1481.0
193.00	1711.0
193.90	406.0
196.00	4587.0
198.00	148032.0
199.00	10050.0
200.00	690.0
201.50	822.0
203.00	910.0
204.00	4768.0
205.00	8353.0
206.00	37352.0
207.00	4957.0
208.00	1135.0
209.00	388.0
211.00	1511.0
215.00	381.0
216.10	741.0
217.00	9081.0
218.00	1247.0
221.00	7246.0
223.00	2163.0
224.00	20216.0
225.00	5471.0
226.00	601.0
227.00	7913.0
228.00	1101.0
229.00	1861.0
231.00	853.0
234.00	539.0
235.00	603.0
236.00	394.0
237.00	792.0
238.90	347.0
240.90	485.0
242.00	1148.0
243.00	1306.0
244.00	17968.0
245.00	2289.0
246.00	3103.0
247.00	602.0
249.00	691.0
252.90	372.0
255.00	92608.0
256.00	13458.0
257.00	1088.0
258.00	4644.0
259.00	807.0
265.00	1917.0
273.00	2946.0
274.00	7796.0
275.00	48016.0
276.00	6513.0
277.00	3554.0
278.00	536.0
283.00	444.0
285.00	633.0
293.00	835.0
296.00	11725.0
297.00	1616.0
303.00	1422.0
304.00	385.0
314.00	628.0
315.00	1329.0
316.00	822.0
321.00	445.0
323.00	4477.0
324.00	766.0
327.00	852.0
327.90	432.0
331.90	338.0
333.00	433.0
334.00	2814.0
335.00	760.0
341.00	529.0
346.00	973.0
352.00	1372.0
353.00	912.0
354.00	1443.0
365.00	5644.0
366.00	880.0
371.00	354.0
372.00	2507.0
373.00	670.0
383.00	688.0
390.00	301.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

402.00	1048.0
403.00	1545.0
404.10	571.0
421.00	1251.0
422.00	1232.0
423.10	8946.0
424.10	2042.0
441.10	23312.0
442.10	149312.0
443.10	28912.0
444.10	2623.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

m/z	Abundance
39.00	1746.0
44.00	363.0
50.00	6109.0
51.00	23656.0
52.00	1133.0
56.00	750.0
57.00	1705.0
60.90	300.0
62.00	356.0
63.00	1002.0
65.00	479.0
69.00	27056.0
74.00	2638.0
75.00	4422.0
76.10	1488.0
77.00	31720.0
78.00	2155.0
79.00	2089.0
80.00	1565.0
81.00	2205.0
82.00	558.0
83.00	556.0
85.00	414.0
86.00	635.0
91.00	551.0
92.00	604.0
93.00	3494.0
98.00	2681.0
99.00	2199.0
101.00	1318.0
103.00	475.0
103.90	839.0
105.00	804.0
107.00	10050.0
108.00	1603.0
110.00	20096.0
111.00	2864.0
112.00	349.0
116.00	568.0
117.00	7653.0
118.00	522.0
122.00	760.0
123.00	1074.0
123.90	448.0
125.00	478.0
127.00	39496.0
128.00	3014.0
129.00	15335.0
130.00	1260.0
134.00	460.0
135.00	1285.0
136.00	518.0
137.00	651.0
141.00	1940.0
142.00	729.0
143.00	432.0
146.00	342.0
147.00	983.0
148.00	2148.0
148.90	471.0
152.90	654.0
154.00	514.0
155.00	1106.0
156.00	1705.0
157.10	330.0
157.90	380.0
159.90	673.0
161.00	909.0
165.00	796.0
166.00	659.0
167.00	4183.0
168.00	2009.0
169.00	329.0
171.90	383.0
173.00	490.0
174.00	953.0
175.00	1619.0
176.00	504.0
177.00	806.0
179.00	3169.0
180.00	2226.0
181.00	1090.0
185.00	1596.0
186.00	12892.0
187.00	3590.0
187.90	413.0
189.00	803.0
191.00	327.0
192.00	1013.0
193.00	1158.0
196.00	3047.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77

198.00	101688.0
199.00	6490.0
200.00	558.0
201.40	510.0
203.00	655.0
204.00	3368.0
205.00	5957.0
206.00	25832.0
207.00	3377.0
208.00	865.0
211.00	1079.0
215.90	501.0
217.00	6649.0
218.00	797.0
221.00	5059.0
223.00	1502.0
224.00	14626.0
225.00	3626.0
226.00	445.0
227.00	5826.0
228.00	879.0
229.00	1170.0
231.00	592.0
234.00	431.0
234.90	479.0
237.00	521.0
240.90	364.0
242.00	805.0
243.10	958.0
244.00	12967.0
245.00	1745.0
246.00	2338.0
247.00	445.0
249.00	490.0
255.00	67440.0
256.00	9694.0
257.00	721.0
258.00	3637.0
259.00	512.0
265.00	1561.0
273.00	2245.0
274.00	5754.0
275.00	34488.0
276.00	4618.0
277.00	2610.0
278.00	397.0
285.00	461.0
293.00	617.0
296.00	8807.0
297.00	1177.0
303.00	1036.0
304.00	317.0
314.00	531.0
315.00	1054.0
316.00	552.0
321.00	310.0
323.00	3266.0
324.00	565.0
327.00	536.0
328.00	343.0
332.90	338.0
334.00	2292.0
335.00	669.0
341.00	476.0
346.00	707.0
352.00	1105.0
353.00	774.0
354.00	1205.0
365.00	4709.0
365.90	686.0
371.00	330.0
372.10	2046.0
373.00	521.0
383.00	543.0
390.00	304.0
402.00	836.0
403.00	1306.0
404.00	420.0
421.00	1024.0
422.10	1061.0
423.10	7820.0
424.10	1649.0
441.10	20136.0
442.10	129136.0
443.10	24408.0
444.10	2091.0

Instrument :
BNA_N
ClientSampleId :
DFTPP77