

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
 Data File : BN023504.D
 Acq On : 28 Dec 2022 02:07
 Operator : CG/JU
 Sample : PB149847BL
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB149847BL

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 12/28/2022
 Supervised By : Jagrut Upadhyay 12/28/2022

Quant Time: Dec 28 04:18:31 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 03:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.977	152	4521	0.400	ng	0.00
7) Naphthalene-d8	10.797	136	12814	0.400	ng	0.00
13) Acenaphthene-d10	14.623	164	7369	0.400	ng	0.00
19) Phenanthrene-d10	17.365	188	15647	0.400	ng	#-0.01
29) Chrysene-d12	21.558	240	12174	0.400	ng	#-0.01
35) Perylene-d12	24.002	264	9765	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.522	112	4513	0.451	ng	0.00
5) Phenol-d6	7.139	99	5489	0.447	ng	0.00
8) Nitrobenzene-d5	9.142	82	3399m	0.339	ng	0.00
11) 2-Methylnaphthalene-d10	12.382	152	6130	0.328	ng	0.00
14) 2,4,6-Tribromophenol	16.123	330	850	0.235	ng	0.00
15) 2-Fluorobiphenyl	13.255	172	10304	0.353	ng	0.00
27) Fluoranthene-d10	19.401	212	13413	0.351	ng	0.00
31) Terphenyl-d14	20.000	244	13827	0.487	ng	0.00

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
Data File : BN023504.D
Acq On : 28 Dec 2022 02:07
Operator : CG/JU
Sample : PB149847BL
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB149847BL

Manual Integrations
APPROVED

Reviewed By :Christian Giraldo 12/28/2022
Supervised By :Jagrut Upadhyay 12/28/2022

Quant Time: Dec 28 04:18:31 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 03:57:20 2022
Response via : Initial Calibration

