

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
 Data File : BN023511.D
 Acq On : 28 Dec 2022 06:20
 Operator : CG/JU
 Sample : N6100-03
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SP-303-20221215

Quant Time: Dec 28 07:15:57 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 03:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.977	152	4662	0.400	ng	0.00
7) Naphthalene-d8	10.786	136	13163	0.400	ng	#-0.01
13) Acenaphthene-d10	14.623	164	7343	0.400	ng	0.00
19) Phenanthrene-d10	17.365	188	15971	0.400	ng	#-0.01
29) Chrysene-d12	21.562	240	10835	0.400	ng	0.00
35) Perylene-d12	24.001	264	8484	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.522	112	1496	0.145	ng	0.00
5) Phenol-d6	7.139	99	1091	0.086	ng	0.00
8) Nitrobenzene-d5	9.142	82	3123	0.303	ng	0.00
11) 2-Methylnaphthalene-d10	12.378	152	6158	0.321	ng	0.00
14) 2,4,6-Tribromophenol	16.111	330	614	0.171	ng	-0.01
15) 2-Fluorobiphenyl	13.244	172	10278	0.353	ng	-0.01
27) Fluoranthene-d10	19.401	212	14555	0.373	ng	0.00
31) Terphenyl-d14	19.996	244	16058	0.636	ng	0.00
Target Compounds						Qvalue
28) Fluoranthene	19.431	202	1182	0.023	ng	# 94
30) Pyrene	19.797	202	963	0.024	ng	96
34) Bis(2-ethylhexyl)phtha...	21.464	149	1704	0.077	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
Data File : BN023511.D
Acq On : 28 Dec 2022 06:20
Operator : CG/JU
Sample : N6100-03
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SP-303-20221215

Quant Time: Dec 28 07:15:57 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 03:57:20 2022
Response via : Initial Calibration

