

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
 Data File : BN023513.D
 Acq On : 28 Dec 2022 07:32
 Operator : CG/JU
 Sample : N6182-05
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWOW-BR-04-223-244-122122

Quant Time: Dec 28 08:08:44 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 03:57:20 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 12/28/2022
 Supervised By : Jagrut Upadhyay 12/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.977	152	4436	0.400	ng	0.00
7) Naphthalene-d8	10.786	136	12504	0.400	ng	#-0.01
13) Acenaphthene-d10	14.623	164	6912	0.400	ng	0.00
19) Phenanthrene-d10	17.365	188	15588	0.400	ng	#-0.01
29) Chrysene-d12	21.562	240	10727	0.400	ng	0.00
35) Perylene-d12	24.001	264	8174	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.522	112	1972	0.201	ng	0.00
5) Phenol-d6	7.139	99	1408	0.117	ng	0.00
8) Nitrobenzene-d5	9.142	82	4241m	0.433	ng	0.00
11) 2-Methylnaphthalene-d10	12.378	152	7886	0.433	ng	0.00
14) 2,4,6-Tribromophenol	16.111	330	1459	0.431	ng	-0.01
15) 2-Fluorobiphenyl	13.244	172	13128	0.479	ng	-0.01
27) Fluoranthene-d10	19.401	212	16909	0.444	ng	0.00
31) Terphenyl-d14	19.996	244	17460	0.698	ng	0.00
Target Compounds						
25) Phenanthrene	17.414	178	1771	0.040	ng	# 95
28) Fluoranthene	19.431	202	1890	0.038	ng	# 93
30) Pyrene	19.793	202	1021	0.025	ng	# 93
34) Bis(2-ethylhexyl)phtha...	21.464	149	2003	0.091	ng	# 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
Data File : BN023513.D
Acq On : 28 Dec 2022 07:32
Operator : CG/JU
Sample : N6182-05
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWOW-BR-04-223-244-122122

Quant Time: Dec 28 08:08:44 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 03:57:20 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 12/28/2022
Supervised By : Jagrut Upadhyay 12/28/2022

