

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
 Data File : BN023516.D
 Acq On : 28 Dec 2022 10:07
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Dec 28 10:54:19 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 03:57:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	121	0.00
2	1,4-Dioxane	0.400	0.404	-1.0	122	0.00
3	n-Nitrosodimethylamine	0.400	0.379	5.3	115	0.00
4 S	2-Fluorophenol	0.400	0.377	5.8	113	0.00
5 S	Phenol-d6	0.400	0.372	7.0	115	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	121	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	124	0.00
8 S	Nitrobenzene-d5	0.400	0.365	8.8	117	0.00
9	Naphthalene	0.400	0.392	2.0	124	0.00
10	Hexachlorobutadiene	0.400	0.402	-0.5	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.359	10.3	111	0.00
12	2-Methylnaphthalene	0.400	0.387	3.3	123	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	122	0.00
14 S	2,4,6-Tribromophenol	0.400	0.328	18.0	111	0.00
15 S	2-Fluorobiphenyl	0.400	0.389	2.8	121	0.00
16	Acenaphthylene	0.400	0.364	9.0	123	0.00
17	Acenaphthene	0.400	0.390	2.5	124	0.00
18	Fluorene	0.400	0.379	5.3	121	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	120	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.479	-19.7	120	0.00
21	4-Bromophenyl-phenylether	0.400	0.384	4.0	121	0.00
22	Hexachlorobenzene	0.400	0.405	-1.3	125	0.00
23	Atrazine	0.400	0.333	16.8	108	0.00
24	Pentachlorophenol	0.400	0.395	1.3	111	0.00
25	Phenanthrene	0.400	0.386	3.5	120	0.00
26	Anthracene	0.400	0.359	10.3	117	0.00
27 SURR	Fluoranthene-d10	0.400	0.360	10.0	114	0.00
28	Fluoranthene	0.400	0.366	8.5	115	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.438	-9.5	115	0.00
31 S	Terphenyl-d14	0.400	0.466	-16.5	125	0.00
32	Benzo(a)anthracene	0.400	0.369	7.8	101	0.00
33	Chrysene	0.400	0.388	3.0	104	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.344	14.0	95	0.00
35 I	Perylene-d12	0.400	0.400	0.0	82	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.389	2.8	84	0.00
37	Benzo(b)fluoranthene	0.400	0.424	-6.0	90	0.00
38	Benzo(k)fluoranthene	0.400	0.403	-0.8	87	0.00
39 C	Benzo(a)pyrene	0.400	0.409	-2.2	87	0.00
40	Dibenzo(a,h)anthracene	0.400	0.381	4.8	83	0.00
41	Benzo(g,h,i)perylene	0.400	0.384	4.0	82	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4