

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
 Data File : BN023487.D
 Acq On : 27 Dec 2022 13:59
 Operator : CG/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Dec 28 03:34:32 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 28 03:33:25 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.977	152	5782	0.400	ng	0.00
7) Naphthalene-d8	10.797	136	15957	0.400	ng	0.00
13) Acenaphthene-d10	14.624	164	9017	0.400	ng	0.00
19) Phenanthrene-d10	17.377	188	18988	0.400	ng	0.00
29) Chrysene-d12	21.563	240	15454	0.400	ng	# 0.00
35) Perylene-d12	24.009	264	12386	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.522	112	2574	0.160	ng	0.00
5) Phenol-d6	7.139	99	3118	0.163	ng	0.00
8) Nitrobenzene-d5	9.142	82	2350	0.182	ng	0.00
11) 2-Methylnaphthalene-d10	12.386	152	4188	0.153	ng	0.00
14) 2,4,6-Tribromophenol	16.124	330	794	0.234	ng	0.00
15) 2-Fluorobiphenyl	13.255	172	7068	0.190	ng	0.00
27) Fluoranthene-d10	19.406	212	8718	0.155	ng	0.00
31) Terphenyl-d14	20.000	244	6545	0.177	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.369	88	1166	0.154	ng	94
3) n-Nitrosodimethylamine	3.702	42	1330	0.193	ng	# 99
6) bis(2-Chloroethyl)ether	7.399	93	3082	0.187	ng	100
9) Naphthalene	10.840	128	7943	0.167	ng	98
10) Hexachlorobutadiene	11.128	225	1620	0.187	ng	# 98
12) 2-Methylnaphthalene	12.458	142	5769	0.182	ng	98
16) Acenaphthylene	14.346	152	6669	0.154	ng	99
17) Acenaphthene	14.688	154	4826	0.164	ng	99
18) Fluorene	15.671	166	6579	0.171	ng	97
20) 4,6-Dinitro-2-methylph...	15.751	198	428	0.176	ng	# 80
21) 4-Bromophenyl-phenylether	16.558	248	2147	0.194	ng	97
22) Hexachlorobenzene	16.682	284	2609	0.213	ng	100
23) Atrazine	16.831	200	1593	0.183	ng	99
24) Pentachlorophenol	17.030	266	769	0.267	ng	98
25) Phenanthrene	17.415	178	10360	0.162	ng	99
26) Anthracene	17.501	178	8136	0.147	ng	99
28) Fluoranthene	19.435	202	11198	0.160	ng	100
30) Pyrene	19.802	202	11459	0.175	ng	100
32) Benzo(a)anthracene	21.545	228	9775	0.177	ng	98
33) Chrysene	21.607	228	10090	0.165	ng	99
34) Bis(2-ethylhexyl)phtha...	21.473	149	6473	0.208	ng	99
36) Indeno(1,2,3-cd)pyrene	26.550	276	8519	0.154	ng	99
37) Benzo(b)fluoranthene	23.258	252	9063	0.190	ng	# 91
38) Benzo(k)fluoranthene	23.308	252	8657	0.170	ng	# 90
39) Benzo(a)pyrene	23.898	252	6686	0.162	ng	# 89
40) Dibenzo(a,h)anthracene	26.574	278	6442	0.146	ng	94
41) Benzo(g,h,i)perylene	27.342	276	7629	0.158	ng	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122722\
Data File : BN023487.D
Acq On : 27 Dec 2022 13:59
Operator : CG/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Dec 28 03:34:32 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 28 03:33:25 2022
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.977 min Scan# 61
 Delta R.T. 0.000 min
 Lab File: BN023487.D
 Acq: 27 Dec 2022 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:152 Resp: 5782
 Ion Ratio Lower Upper
 152 100
 150 159.8 125.0 187.6
 115 62.0 48.8 73.2



#2
 1,4-Dioxane
 Concen: 0.154 ng
 RT: 3.369 min Scan# 40
 Delta R.T. 0.000 min
 Lab File: BN023487.D
 Acq: 27 Dec 2022 13:59

Tgt Ion: 88 Resp: 1166
 Ion Ratio Lower Upper
 88 100
 43 35.8 24.6 36.8
 58 68.4 58.2 87.4









#3
n-Nitrosodimethylamine
Concen: 0.193 ng
RT: 3.702 min Scan# 86
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 1330
Ion Ratio Lower Upper
42 100
74 128.5 102.1 153.1
44 9.6 5.6 8.4#



#4
2-Fluorophenol
Concen: 0.160 ng
RT: 5.522 min Scan# 338
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion: 112 Resp: 2574
Ion Ratio Lower Upper
112 100
64 56.4 44.3 66.5
63 29.4 23.8 35.6









#5
Phenol-d6
Concen: 0.163 ng
RT: 7.139 min Scan# 562
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	99	Resp:	3118
Ion	Ratio	Lower	Upper
99	100		
42	21.3	16.8	25.2
71	33.3	27.0	40.4



#6
bis(2-Chloroethyl)ether
Concen: 0.187 ng
RT: 7.399 min Scan# 598
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:	93	Resp:	3082
Ion	Ratio	Lower	Upper
93	100		
63	73.8	58.9	88.3
95	32.4	26.1	39.1









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.797 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	15957
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	7.3	5.8	8.8
68	6.3	5.2	7.8



#8
Nitrobenzene-d5
Concen: 0.182 ng
RT: 9.142 min Scan# 802
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:	82	Resp:	2350
Ion	Ratio	Lower	Upper
82	100		
128	42.9	32.6	49.0
54	55.1	41.4	62.0









#9
Naphthalene
Concen: 0.167 ng
RT: 10.840 min Scan# 961
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	128	Resp:	7943
Ion	Ratio	Lower	Upper
128	100		
129	11.9	9.1	13.7
127	14.5	11.0	16.4



#10
Hexachlorobutadiene
Concen: 0.187 ng
RT: 11.128 min Scan# 988
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:	225	Resp:	1620
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.2	51.8	77.8









#11
2-Methylnaphthalene-d10
Concen: 0.153 ng
RT: 12.386 min Scan# 11
Delta R.T. 0.004 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 4188
Ion Ratio Lower Upper
152 100
151 20.7 14.8 22.2



#12
2-Methylnaphthalene
Concen: 0.182 ng
RT: 12.458 min Scan# 1274
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:142 Resp: 5769
Ion Ratio Lower Upper
142 100
141 89.8 70.5 105.7
115 39.2 29.5 44.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.624 min Scan# 1485
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 9017
Ion Ratio Lower Upper
164 100
162 105.0 83.9 125.9
160 49.7 40.9 61.3



#14
2,4,6-Tribromophenol
Concen: 0.234 ng
RT: 16.124 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:330 Resp: 794
Ion Ratio Lower Upper
330 100
332 96.5 76.7 115.1
141 36.1 29.8 44.6









#15
2-Fluorobiphenyl
Concen: 0.190 ng
RT: 13.255 min Scan# 1357
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172 Resp: 7068
Ion Ratio Lower Upper
172 100
171 35.2 27.6 41.4
170 23.6 18.7 28.1



#16
Acenaphthylene
Concen: 0.154 ng
RT: 14.346 min Scan# 1459
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:152 Resp: 6669
Ion Ratio Lower Upper
152 100
151 19.8 15.4 23.2
153 13.3 10.3 15.5









#17
Acenaphthene
Concen: 0.164 ng
RT: 14.688 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:154 Resp: 4826
Ion Ratio Lower Upper
154 100
153 114.1 90.8 136.2
152 57.5 46.3 69.5



#18
Fluorene
Concen: 0.171 ng
RT: 15.671 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:166 Resp: 6579
Ion Ratio Lower Upper
166 100
165 101.7 79.0 118.6
167 12.7 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.377 min Scan# 1721

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 18988

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.1 6.7 10.1



#20

4,6-Dinitro-2-methylphenol

Concen: 0.176 ng

RT: 15.751 min Scan# 1590

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59




Tgt Ion:198 Resp: 428


Ion Ratio Lower Upper

198 100

51 79.9 48.2 72.4#

105 60.5 41.0 61.4









#21
4-Bromophenyl-phenylether
Concen: 0.194 ng
RT: 16.558 min Scan# 1655
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:248 Resp: 2147
Ion Ratio Lower Upper
248 100
250 91.7 72.6 108.8
141 82.4 69.4 104.0



#22
Hexachlorobenzene
Concen: 0.213 ng
RT: 16.682 min Scan# 1665
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:284 Resp: 2609
Ion Ratio Lower Upper
284 100
142 36.3 29.0 43.6
249 30.5 24.5 36.7





#23

Atrazine

Concen: 0.183 ng

RT: 16.831 min Scan# 1677

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:200 Resp: 1593

Ion Ratio Lower Upper

200 100

173 28.1 22.1 33.1

215 48.3 38.2 57.2



#24

Pentachlorophenol

Concen: 0.267 ng

RT: 17.030 min Scan# 1693

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59




Tgt Ion:266 Resp: 769


Ion Ratio Lower Upper

266 100

264 60.5 49.7 74.5

268 62.2 51.4 77.0









#25
Phenanthrene
Concen: 0.162 ng
RT: 17.415 min Scan# 1724
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 10360
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.7 12.2 18.4



#26
Anthracene
Concen: 0.147 ng
RT: 17.501 min Scan# 1731
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:178 Resp: 8136
Ion Ratio Lower Upper
178 100
176 19.1 15.0 22.6
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.155 ng
RT: 19.406 min Scan# 2103
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:212 Resp: 8718
Ion Ratio Lower Upper
212 100
106 15.3 12.3 18.5
104 8.9 6.8 10.2



#28
Fluoranthene
Concen: 0.160 ng
RT: 19.435 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:202 Resp: 11198
Ion Ratio Lower Upper
202 100
101 12.1 9.6 14.4
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.563 min Scan# 2419
Delta R.T. -0.009 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:240 Resp: 15454
Ion Ratio Lower Upper
240 100
120 19.1 8.9 13.3#
236 28.9 22.5 33.7



#30
Pyrene
Concen: 0.175 ng
RT: 19.802 min Scan# 2197
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:202 Resp: 11459
Ion Ratio Lower Upper
202 100
200 21.3 17.0 25.4
203 18.2 14.4 21.6









#31
 Terphenyl-d14
 Concen: 0.177 ng
 RT: 20.000 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: BN023487.D
 Acq: 27 Dec 2022 13:59




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2


Tgt Ion:244 Resp: 6545
 Ion Ratio Lower Upper
 244 100
 212 10.1 7.4 11.0
 122 0.0 0.0 0.0



#32
 Benzo(a)anthracene
 Concen: 0.177 ng
 RT: 21.545 min Scan# 2417
 Delta R.T. -0.009 min
 Lab File: BN023487.D
 Acq: 27 Dec 2022 13:59

Tgt Ion:228 Resp: 9775
 Ion Ratio Lower Upper
 228 100
 226 28.9 21.8 32.8
 229 19.9 16.0 24.0





#33

Chrysene

Concen: 0.165 ng

RT: 21.607 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 10090

Ion Ratio Lower Upper

228 100

226 29.9 24.1 36.1

229 20.4 16.0 24.0



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.208 ng

RT: 21.473 min Scan# 2409

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59




Tgt Ion:149 Resp: 6473


Ion Ratio Lower Upper

149 100

167 26.5 21.0 31.6

279 3.3 2.2 3.4









#35
Perylene-d12
Concen: 0.400 ng
RT: 24.009 min Scan# 3062
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	264	Resp:	12386
Ion Ratio	Lower	Upper	
264	100		
260	26.4	21.1	31.7
265	60.6	48.9	73.3



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.154 ng
RT: 26.550 min Scan# 3931
Delta R.T. -0.006 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Tgt Ion:	276	Resp:	8519
Ion Ratio	Lower	Upper	
276	100		
138	27.9	23.0	34.6
277	24.7	19.5	29.3





#37

Benzo(b)fluoranthene

Concen: 0.190 ng

RT: 23.258 min Scan# 2805

Delta R.T. -0.003 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 9063

Ion Ratio Lower Upper

252 100

253 29.2 20.7 31.1

125 21.9 13.4 20.2#



#38

Benzo(k)fluoranthene

Concen: 0.170 ng

RT: 23.308 min Scan# 2822

Delta R.T. -0.003 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59




Tgt Ion:252 Resp: 8657


Ion Ratio Lower Upper

252 100

253 29.6 20.2 30.4

125 21.5 13.0 19.6#





#39

Benzo(a)pyrene

Concen: 0.162 ng

RT: 23.898 min Scan# 3024

Delta R.T. 0.000 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 6686

Ion Ratio Lower Upper

252 100

253 33.2 22.3 33.5

125 29.9 18.7 28.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.146 ng

RT: 26.574 min Scan# 3939

Delta R.T. -0.003 min

Lab File: BN023487.D

Acq: 27 Dec 2022 13:59




Tgt Ion:278 Resp: 6442


Ion Ratio Lower Upper

278 100

139 24.6 17.7 26.5


279 31.3 22.2 33.2





#41
Benzo(g,h,i)perylene
Concen: 0.158 ng
RT: 27.342 min Scan# 41
Delta R.T. 0.000 min
Lab File: BN023487.D
Acq: 27 Dec 2022 13:59

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2



Tgt Ion:276 Resp: 7629
Ion Ratio Lower Upper
276 100
277 26.2 20.0 30.0
138 29.2 20.6 31.0

