

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122819\
 Data File : BN009231.D
 Acq On : 28 Dec 2019 16:54
 Operator : JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDC0.426

Manual Integrations
 APPROVED

mohammad
 12/29/2019 5:32:39 PM

Quant Time: Dec 29 04:40:12 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN121619.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Dec 29 03:34:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.51	152	2755	0.40	ng/ul	0.00
2) Naphthalene-d8	10.27	136	10075	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.14	164	5247	0.40	ng/ul	0.00
10) Phenanthrene-d10	16.89	188	10648	0.40	ng/ul	0.00
16) Chrysene-d12	21.09	240	9526	0.40	ng/ul	0.00
20) Perylene-d12	23.22	264	10054	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.86	152	6500	0.36	ng/ul	0.00
14) Fluoranthene-d10	18.93	212	13699	0.44	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.31	128	12250	0.401	ng/ul	100
5) 2-Methylnaphthalene	11.93	142	8385	0.363	ng/ul	100
7) Acenaphthylene	13.86	152	11116	0.403	ng/ul	99
8) Acenaphthene	14.20	153	8174	0.378	ng/ul	99
9) Fluorene	15.20	166	9129	0.357	ng/ul	99
11) Pentachlorophenol	16.56	266	1469	0.518	ng/ul	100
12) Phenanthrene	16.93	178	14661	0.396	ng/ul	100
13) Anthracene	17.02	178	14023	0.422	ng/ul	99
15) Fluoranthene	18.95	202	17904	0.423	ng/ul	98
17) Pyrene	19.32	202	18366	0.378	ng/ul	99
18) Benzo(a)anthracene	21.07	228	17944	0.424	ng/ul	99
19) Chrysene	21.12	228	17381	0.409	ng/ul	99
21) Benzo(b)fluoranthene	22.59	252	17611	0.375	ng/ul	99
22) Benzo(k)fluoranthene	22.63	252	18253m	0.386	ng/ul	
23) Benzo(a)pyrene	23.12	252	17088	0.415	ng/ul	98
24) Indeno(1,2,3-cd)pyrene	25.31	276	19941	0.445	ng/ul	96
25) Dibenzo(a,h)anthracene	25.32	278	16157	0.453	ng/ul	99
26) Benzo(g,h,i)perylene	25.94	276	16916	0.448	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN122819\
Data File : BN009231.D
Acq On : 28 Dec 2019 16:54
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.426

Manual Integrations
APPROVED

mohammad
12/29/2019 5:32:39 PM

Quant Time: Dec 29 04:40:12 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN121619.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Dec 29 03:34:38 2019
Response via : Initial Calibration

