


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122819\
Data File : BN009198.D
Acq On : 27 Dec 2019 17:16
Operator : JU
Sample : DFTPP18
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP18

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN121619.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Mon Dec 30 01:21:37 2019



AutoFind: Scans 2443, 2444, 2445; Background Corrected with Scan 2437

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.1	101474	PASS
68	69	0.00	2	2.0	1742	PASS
69	198	0.00	100	41.4	85528	PASS
70	69	0.00	2	0.7	609	PASS
127	198	10	80	46.2	95426	PASS
197	198	0.00	2	0.2	487	PASS
198	198	100	100	100.0	206656	PASS
199	198	5	9	6.4	13281	PASS
275	198	10	60	25.8	53346	PASS
365	198	1	100	3.4	6943	PASS
441	198	0.01	100	11.1	22954	PASS
442	198	50	100	70.2	145029	PASS
443	442	15	24	19.2	27858	PASS

m/z	Abundance
39.10	585.0
40.00	2032.0
41.00	665.0
42.10	407.0
43.10	1207.0
44.00	5470.0
45.00	425.0
49.00	474.0
50.90	590.0
52.80	301.0
54.90	584.0
57.00	722.0
59.00	306.0
60.10	603.0
69.00	373.0
73.10	1194.0
84.00	366.0
91.00	327.0
132.90	392.0
193.10	324.0
207.00	2150.0
208.10	720.0
208.80	329.0
281.00	601.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
37.00	539.0
38.00	1378.0
39.00	8316.0
40.00	2088.0
41.00	734.0
41.90	502.0
43.00	1058.0
44.00	5765.0
45.10	842.0
46.90	424.0
49.00	1532.0
50.00	24520.0
51.00	100264.0
52.10	5640.0
53.00	433.0
54.90	1067.0
56.00	3165.0
57.00	8682.0
58.00	516.0
59.00	315.0
60.00	312.0
61.00	1422.0
62.00	1205.0
63.10	3419.0
64.00	693.0
65.00	1587.0
67.10	374.0
68.10	1904.0
69.00	84352.0
70.00	606.0
73.00	2055.0
74.00	7843.0
75.00	13093.0
76.10	4155.0
77.00	93080.0
78.00	6692.0
79.00	6350.0
80.00	4579.0
81.00	7792.0
81.90	1309.0
83.00	1839.0
83.90	588.0
85.00	1599.0
86.00	2116.0
86.90	718.0
87.80	341.0
88.10	359.0
91.00	1710.0
92.00	2116.0
93.00	10341.0
94.00	799.0
96.00	787.0
96.90	481.0
98.00	7198.0
98.90	7267.0
99.90	569.0
101.00	3600.0
101.90	317.0
102.90	1418.0
104.00	2350.0
105.00	2237.0
106.10	517.0
107.00	28424.0
108.00	4061.0
108.90	884.0
110.00	46152.0
111.00	6514.0
111.90	944.0
112.90	436.0
114.90	320.0
116.00	1177.0
117.00	21496.0
118.00	1487.0
118.80	347.0
119.90	391.0
122.00	1719.0
123.00	3019.0
124.00	1444.0
125.00	1296.0
125.90	423.0
127.00	88208.0
128.00	7206.0
129.00	33144.0
129.90	2663.0
131.00	645.0
132.00	407.0
134.00	1116.0
135.00	3020.0
136.00	1240.0
137.00	1680.0
140.00	456.0
141.00	4239.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

142.10	1200.0
142.90	1376.0
143.80	393.0
145.90	887.0
147.00	2973.0
148.00	5394.0
149.00	1318.0
150.10	403.0
151.10	571.0
153.00	1641.0
153.90	1105.0
155.00	3109.0
156.00	3630.0
157.00	962.0
158.00	705.0
158.90	834.0
159.90	1301.0
161.00	2549.0
162.10	608.0
163.90	432.0
165.00	1518.0
166.00	1417.0
167.00	8585.0
168.00	6040.0
168.90	826.0
169.90	484.0
170.90	453.0
171.70	815.0
172.90	890.0
173.90	1885.0
175.00	3255.0
176.00	975.0
176.90	1611.0
178.10	789.0
178.90	6821.0
180.00	4444.0
181.00	1823.0
182.00	562.0
184.00	505.0
185.00	3415.0
186.00	25712.0
187.00	7543.0
188.00	674.0
188.90	1614.0
190.80	536.0
192.00	1986.0
193.00	2521.0
193.90	622.0
195.90	4133.0
197.90	167424.0
198.90	10836.0
200.00	1026.0
200.70	314.0
201.30	821.0
201.70	510.0
202.90	1532.0
204.10	5471.0
205.00	10519.0
206.00	46112.0
207.10	7219.0
208.00	1942.0
208.90	755.0
210.00	621.0
210.90	1792.0
212.10	420.0
214.90	698.0
216.00	1062.0
217.00	11633.0
218.10	1419.0
220.30	343.0
221.10	6609.0
223.10	2388.0
224.00	23056.0
225.00	6934.0
226.00	815.0
227.00	10071.0
228.00	1368.0
229.00	1666.0
229.80	392.0
230.00	376.0
231.00	504.0
233.90	629.0
234.90	654.0
235.90	432.0
237.00	1275.0
238.00	300.0
239.00	324.0
240.10	442.0
240.90	850.0
242.00	1264.0
243.00	1456.0
244.00	17192.0
245.00	2197.0
245.90	3161.0
246.90	732.0
248.90	867.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

254.00	852.0
255.00	90008.0
256.00	12571.0
256.90	917.0
258.00	4492.0
258.90	807.0
264.90	1650.0
270.90	358.0
273.00	2141.0
274.00	7137.0
275.00	40344.0
275.90	5278.0
277.00	2881.0
277.90	404.0
281.00	570.0
282.00	315.0
283.00	632.0
285.10	384.0
289.10	436.0
292.90	842.0
296.00	10128.0
296.90	1375.0
303.00	1256.0
303.90	492.0
314.00	536.0
314.90	1144.0
316.10	1097.0
320.90	523.0
323.00	3429.0
324.00	588.0
327.10	608.0
328.00	365.0
334.00	1805.0
335.10	820.0
340.80	749.0
346.00	866.0
351.90	1143.0
352.90	783.0
354.00	1115.0
365.00	4219.0
366.00	852.0
372.00	1643.0
373.10	371.0
383.10	334.0
401.90	640.0
403.00	1082.0
403.90	442.0
420.90	851.0
422.10	1011.0
423.00	4992.0
424.00	1887.0
441.00	13120.0
442.00	85392.0
443.00	16568.0
444.10	1600.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
37.10	665.0
38.00	1246.0
39.10	10607.0
40.00	2500.0
41.10	1392.0
42.10	475.0
43.00	1305.0
44.00	5432.0
45.00	792.0
49.10	1622.0
50.00	32896.0
51.00	123376.0
52.00	6873.0
53.20	409.0
55.00	1090.0
56.00	3383.0
57.00	8167.0
57.80	360.0
60.00	498.0
61.00	1760.0
62.00	2035.0
63.00	5341.0
64.00	700.0
65.10	2518.0
68.00	2062.0
69.00	103384.0
70.00	863.0
71.00	317.0
73.00	2259.0
74.00	10470.0
75.00	16222.0
76.00	6371.0
77.00	118592.0
78.10	8711.0
79.00	8094.0
80.00	7108.0
81.00	8151.0
82.00	1809.0
83.10	2228.0
83.90	602.0
85.00	1864.0
86.00	3466.0
87.00	1368.0
87.90	535.0
91.00	2428.0
92.00	2385.0
93.00	12225.0
93.90	729.0
95.10	531.0
96.00	985.0
97.20	367.0
98.00	9648.0
99.00	8660.0
100.00	554.0
100.90	4503.0
101.90	382.0
103.10	1366.0
104.00	2717.0
105.00	2672.0
106.10	761.0
107.00	34176.0
108.00	5394.0
109.00	1021.0
110.00	63528.0
111.00	8382.0
111.90	1287.0
112.90	307.0
115.00	347.0
116.00	2258.0
117.00	25648.0
118.00	1962.0
119.10	442.0
120.00	467.0
122.00	2095.0
123.00	3956.0
124.00	1669.0
125.10	1838.0
127.00	117616.0
128.00	9426.0
129.00	45784.0
130.00	3930.0
131.00	811.0
132.00	663.0
132.80	426.0
133.90	1330.0
134.90	3738.0
136.00	1819.0
137.00	2155.0
137.80	390.0
139.90	846.0
141.00	6595.0
142.00	1667.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

143.00	1828.0
144.90	461.0
145.90	992.0
147.00	3080.0
148.00	7768.0
149.00	1409.0
150.00	561.0
150.90	844.0
151.90	384.0
152.90	1917.0
153.90	1332.0
155.00	3705.0
156.00	4944.0
157.00	1108.0
158.00	1298.0
159.00	798.0
159.90	1727.0
161.10	3166.0
161.90	1163.0
162.90	309.0
165.00	2795.0
166.10	1407.0
167.00	11640.0
168.00	6630.0
169.00	1384.0
169.90	319.0
171.00	476.0
172.00	1207.0
173.00	893.0
174.00	2772.0
175.00	3988.0
176.00	1981.0
176.90	1955.0
177.90	830.0
179.00	9418.0
180.00	6404.0
180.90	2635.0
182.10	623.0
183.00	416.0
184.10	872.0
185.00	3937.0
186.00	34368.0
187.00	9057.0
188.00	1355.0
189.00	2303.0
189.80	372.0
191.00	1223.0
192.00	2361.0
193.00	3641.0
194.00	701.0
194.70	332.0
196.00	6376.0
197.10	1463.0
197.90	260480.0
199.00	17272.0
200.00	1440.0
201.40	1327.0
203.00	2085.0
204.00	8767.0
205.00	14483.0
206.00	62400.0
207.00	10680.0
208.00	2350.0
209.00	1394.0
210.00	1481.0
211.00	1972.0
214.90	704.0
216.00	1152.0
217.00	17640.0
217.90	2708.0
221.00	11111.0
222.10	1570.0
223.00	4766.0
224.00	35128.0
225.00	9625.0
226.00	1164.0
227.00	14771.0
228.00	2252.0
229.00	3058.0
229.90	386.0
231.00	1458.0
231.90	382.0
233.90	1306.0
235.00	1099.0
236.00	571.0
236.90	1874.0
238.90	400.0
240.00	499.0
241.00	1165.0
242.00	1560.0
243.00	2700.0
244.00	25320.0
245.00	3531.0
246.00	5784.0
247.00	1097.0
247.90	340.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

249.00	945.0
250.90	439.0
253.00	651.0
255.00	139392.0
256.00	21008.0
256.90	1905.0
258.00	8198.0
259.00	720.0
260.00	446.0
264.90	3918.0
265.80	505.0
270.90	315.0
271.80	414.0
273.00	4476.0
274.00	10093.0
275.00	65088.0
276.00	8342.0
277.00	5206.0
278.00	997.0
281.00	627.0
282.00	440.0
283.10	819.0
284.00	478.0
285.00	690.0
293.00	1576.0
295.10	614.0
296.00	16864.0
297.00	2012.0
303.00	1602.0
304.00	512.0
314.00	940.0
315.00	1874.0
316.00	993.0
321.10	480.0
323.10	5120.0
324.10	1093.0
326.90	1129.0
327.90	754.0
331.90	415.0
333.10	583.0
334.00	3780.0
335.10	890.0
340.90	338.0
341.10	334.0
345.90	857.0
351.90	1288.0
353.00	1262.0
354.00	1227.0
355.00	573.0
365.00	8044.0
365.90	1151.0
371.00	388.0
372.00	2905.0
373.00	727.0
382.90	1044.0
402.00	1270.0
402.90	1814.0
403.80	823.0
421.00	1336.0
422.00	1317.0
423.00	11340.0
424.00	2905.0
424.90	551.0
441.00	28768.0
442.00	172096.0
443.00	34448.0
443.90	3181.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

m/z	Abundance
37.00	384.0
38.00	832.0
39.00	6293.0
40.10	2289.0
41.00	877.0
41.90	502.0
43.00	1116.0
44.00	5160.0
45.00	526.0
49.00	1432.0
50.00	18800.0
51.00	82552.0
52.10	3918.0
53.00	564.0
55.00	1122.0
56.00	2538.0
57.00	6220.0
58.00	746.0
59.00	361.0
60.00	657.0
60.90	1291.0
62.00	943.0
63.00	3240.0
63.80	488.0
65.10	1597.0
67.30	306.0
68.10	1261.0
69.00	69968.0
70.10	359.0
73.00	2512.0
74.00	6841.0
75.00	11371.0
76.00	3898.0
77.00	80136.0
78.00	6179.0
79.00	6028.0
80.00	3777.0
81.00	5157.0
82.00	1446.0
83.00	1237.0
83.70	384.0
84.10	413.0
85.10	995.0
86.00	1965.0
87.00	642.0
88.00	360.0
88.90	358.0
90.90	1384.0
92.00	1639.0
93.00	9374.0
93.90	595.0
95.00	339.0
96.00	889.0
98.00	7104.0
99.00	6540.0
100.00	647.0
101.00	3169.0
103.00	1351.0
104.00	1850.0
104.90	2227.0
105.90	591.0
107.00	23248.0
108.00	3346.0
109.10	732.0
110.00	41160.0
111.00	5691.0
112.00	1073.0
112.90	331.0
115.00	404.0
116.00	1062.0
117.00	18008.0
118.00	1517.0
119.00	315.0
119.80	450.0
122.00	1729.0
123.00	2733.0
123.90	1326.0
125.00	918.0
127.00	80456.0
128.00	6116.0
129.00	31152.0
130.00	2318.0
130.90	652.0
132.90	412.0
134.00	1175.0
134.90	2946.0
136.00	1041.0
137.10	1976.0
138.00	419.0
140.00	791.0
141.00	4365.0
142.00	1113.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

143.00	731.0
143.90	374.0
145.90	988.0
147.00	2754.0
147.90	5655.0
149.00	800.0
149.90	486.0
151.00	703.0
151.70	363.0
152.90	1386.0
154.00	1320.0
155.00	1920.0
156.00	4191.0
157.00	478.0
157.70	857.0
159.00	818.0
159.90	1336.0
160.90	1744.0
161.90	774.0
164.90	2375.0
166.00	897.0
167.00	9457.0
168.00	4223.0
168.90	692.0
170.00	359.0
170.80	476.0
171.90	770.0
173.00	1116.0
174.00	1531.0
175.00	3630.0
176.00	1031.0
176.90	1847.0
177.90	573.0
179.00	7255.0
179.90	4656.0
181.00	2586.0
182.00	503.0
182.90	439.0
183.80	358.0
184.90	2452.0
186.00	26288.0
187.00	7936.0
188.10	1026.0
188.90	1614.0
189.90	416.0
190.90	1661.0
192.00	2304.0
193.00	2790.0
194.00	630.0
195.00	328.0
196.00	5567.0
196.80	992.0
198.00	192064.0
198.90	11736.0
200.00	1027.0
201.40	940.0
202.90	1194.0
204.00	6799.0
205.00	11781.0
206.00	44440.0
207.00	7584.0
208.00	2482.0
209.00	639.0
210.20	793.0
211.00	1995.0
212.00	666.0
214.90	673.0
216.00	1160.0
216.90	12831.0
217.90	1244.0
219.00	316.0
220.20	646.0
221.00	8869.0
221.70	1965.0
222.90	3624.0
224.00	30176.0
225.00	6804.0
226.10	1023.0
227.00	10930.0
228.00	1532.0
228.90	2300.0
231.00	1192.0
233.90	914.0
234.90	776.0
236.00	527.0
237.00	2048.0
239.80	416.0
240.90	704.0
241.90	1251.0
243.10	1286.0
244.00	21312.0
245.00	2938.0
246.00	4599.0
247.00	942.0
248.80	873.0
253.00	625.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18

254.10	977.0
255.00	111696.0
256.00	15608.0
257.00	1262.0
258.00	7233.0
258.90	1037.0
261.00	306.0
265.00	2540.0
266.00	519.0
271.00	365.0
273.00	3521.0
274.00	9246.0
275.00	54608.0
276.00	6388.0
277.00	4526.0
278.00	1012.0
281.00	654.0
282.10	439.0
283.00	839.0
285.00	547.0
286.10	314.0
293.00	738.0
296.00	13290.0
297.00	1944.0
301.90	366.0
303.00	1804.0
304.10	306.0
314.00	731.0
314.90	1237.0
315.90	1008.0
321.00	385.0
323.00	4175.0
324.00	571.0
326.90	1121.0
327.90	434.0
333.00	322.0
334.00	2801.0
334.80	736.0
340.90	640.0
345.90	995.0
352.00	1600.0
353.00	1303.0
354.00	1544.0
354.90	473.0
365.00	8568.0
365.90	862.0
372.00	2784.0
373.00	761.0
383.00	690.0
389.80	501.0
402.00	1153.0
402.90	1712.0
404.10	619.0
421.00	1262.0
422.00	1224.0
423.00	9316.0
423.90	2815.0
441.00	26976.0
442.00	177600.0
443.00	32560.0
444.00	3038.0

Instrument :
BNA_N
ClientSampleId :
DFTPP18