

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122823\
 Data File : BN029207.D
 Acq On : 28 Dec 2023 10:49
 Operator : MA/JU
 Sample : 05960-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 COMA6

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/29/2023
 Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 28 11:33:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 28 01:20:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.739	152	786	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	1530	0.400	ng	0.01
13) Acenaphthene-d10	14.383	164	967	0.400	ng	0.01
19) Phenanthrene-d10	17.131	188	1548	0.400	ng	# 0.00
29) Chrysene-d12	21.349	240	540	0.400	ng	0.00
35) Perylene-d12	23.636	264	617	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	211	0.115	ng	0.00
5) Phenol-d6	6.952	99	85	0.041	ng	0.01
8) Nitrobenzene-d5	8.918	82	476	0.248	ng	0.01
11) 2-Methylnaphthalene-d10	12.129	152	679m	0.257	ng	0.00
14) 2,4,6-Tribromophenol	15.890	330	191	0.256	ng	0.01
15) 2-Fluorobiphenyl	13.014	172	1419	0.274	ng	0.01
27) Fluoranthene-d10	19.173	212	886	0.281	ng	0.00
31) Terphenyl-d14	19.787	244	612	0.366	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.277	149	218	0.048	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN122823\
Data File : BN029207.D
Acq On : 28 Dec 2023 10:49
Operator : MA/JU
Sample : 05960-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
COMA6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/29/2023
Supervised By :mohammad ahmed 12/29/2023

Quant Time: Dec 28 11:33:36 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN122723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 28 01:20:47 2023
Response via : Initial Calibration

