

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN123124\
 Data File : BN035868.D
 Acq On : 31 Dec 2024 14:31
 Operator : RC/JU
 Sample : SSTD3.254
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2242

Quant Time: Dec 31 15:04:59 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 31 14:31:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.837	152	1313	0.400	ng/ul	0.00
4) Naphthalene-d8	10.632	136	2790	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.474	164	1479	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.213	188	2870	0.400	ng/ul	0.00
17) Chrysene-d12	21.390	240	2568	0.400	ng/ul	0.00
23) Perylene-d12	23.699	264	2821	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.288	96	4097	3.070	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.322	88	4532	3.130	ng/ul#	81
14) Pentachlorophenol	16.858	266	2613	3.984	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN123124\
Data File : BN035868.D
Acq On : 31 Dec 2024 14:31
Operator : RC/JU
Sample : SST3.254
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2242

Quant Time: Dec 31 15:04:59 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 31 14:31:37 2024
Response via : Initial Calibration

