



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN123124\
Data File : BN035862.D
Acq On : 31 Dec 2024 10:53
Operator : RC/JU
Sample : DFTPP253
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP253

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN123124.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Dec 31 15:06:22 2024



AutoFind: Scans 2522, 2523, 2524; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.7	48427	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	38.4	50768	PASS
70	69	0.00	2	0.7	361	PASS
127	198	10	80	42.7	56408	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	132043	PASS
199	198	5	9	6.9	9111	PASS
275	198	10	60	26.5	35053	PASS
365	198	1	100	3.3	4419	PASS
441	198	0.01	100	11.1	14621	PASS
442	198	50	100	70.5	93099	PASS
443	442	15	24	20.0	18575	PASS

m/z	Abundance
38.90	172.0
40.00	2358.0
41.00	373.0
42.00	170.0
43.00	411.0
44.00	2939.0
45.00	188.0
54.90	254.0
56.00	156.0
57.00	256.0
73.00	808.0
81.00	158.0
91.10	190.0
135.10	155.0
146.90	355.0
192.90	231.0
207.00	1652.0
208.00	254.0
208.90	291.0
221.00	445.0
281.10	399.0
355.10	189.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
36.90	227.0
38.00	865.0
39.10	4197.0
40.00	2282.0
41.00	437.0
43.10	398.0
44.00	3060.0
44.90	181.0
49.00	629.0
50.00	15551.0
51.00	52000.0
52.00	2581.0
53.00	163.0
55.00	462.0
56.00	1724.0
57.00	3703.0
57.90	288.0
59.00	170.0
60.90	763.0
61.90	806.0
63.00	2335.0
64.00	313.0
65.00	1107.0
69.00	56976.0
70.00	410.0
71.00	153.0
73.00	989.0
74.00	5749.0
75.00	9386.0
76.00	3028.0
77.00	59040.0
78.10	3763.0
79.00	4174.0
80.00	3339.0
81.00	4810.0
82.00	1191.0
83.00	985.0
85.00	844.0
86.00	1280.0
86.90	547.0
87.90	232.0
90.90	1131.0
92.00	1219.0
93.00	7246.0
94.00	644.0
95.10	264.0
96.00	468.0
98.00	5477.0
99.00	3882.0
99.90	361.0
101.00	2303.0
102.90	679.0
104.00	1673.0
105.00	1315.0
106.10	874.0
107.00	17472.0
108.00	2501.0
110.00	31112.0
111.00	5140.0
111.90	475.0
113.00	182.0
115.00	163.0
117.00	14731.0
118.00	1138.0
119.80	216.0
122.00	1058.0
123.00	1807.0
123.90	899.0
125.00	574.0
127.00	55120.0
128.00	4557.0
129.00	22176.0
130.00	2108.0
130.90	474.0
132.10	245.0
133.00	173.0
133.90	707.0
135.00	2162.0
135.90	845.0
137.00	1058.0
140.00	323.0
141.00	2798.0
142.00	1001.0
143.00	564.0
145.00	172.0
146.00	549.0
147.00	2048.0
148.00	3421.0
149.00	644.0
149.90	275.0
151.00	572.0
152.10	227.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

152.90	843.0
153.90	856.0
155.00	1415.0
156.00	2551.0
156.90	504.0
157.90	490.0
159.00	506.0
159.90	960.0
160.90	1261.0
161.90	412.0
162.90	169.0
164.10	191.0
165.00	1228.0
166.00	916.0
167.00	5971.0
168.00	2763.0
169.00	401.0
170.00	210.0
171.00	307.0
172.00	485.0
173.00	703.0
174.00	1011.0
175.00	2484.0
175.90	604.0
177.00	1123.0
178.00	462.0
179.00	4835.0
180.00	2569.0
180.90	1273.0
181.90	183.0
182.90	177.0
184.00	423.0
185.00	2563.0
186.00	18720.0
187.00	5091.0
187.90	557.0
189.00	1041.0
189.90	192.0
191.00	678.0
192.00	1539.0
193.00	1763.0
194.00	450.0
196.00	3109.0
198.00	129280.0
199.00	8898.0
200.00	682.0
201.40	603.0
202.90	841.0
204.00	4668.0
205.10	7750.0
206.00	33888.0
207.00	5225.0
208.00	1315.0
209.00	455.0
210.10	593.0
211.00	1323.0
211.60	476.0
212.10	204.0
214.90	401.0
216.00	671.0
217.00	8751.0
218.00	1241.0
221.00	5281.0
222.00	1294.0
223.00	2316.0
224.00	18600.0
225.00	4572.0
226.10	726.0
227.00	7577.0
227.90	1108.0
229.00	1581.0
230.10	186.0
230.90	534.0
233.80	431.0
235.00	523.0
236.00	472.0
236.90	640.0
238.90	387.0
239.90	275.0
240.90	518.0
242.00	986.0
243.00	977.0
244.00	12992.0
245.00	1647.0
245.90	2750.0
246.90	664.0
249.00	598.0
250.80	172.0
251.90	172.0
253.00	340.0
255.00	74016.0
256.00	9771.0
256.90	878.0
258.00	4096.0
258.90	621.0
259.90	231.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

264.90	1419.0
265.90	276.0
267.10	249.0
271.00	287.0
271.90	294.0
273.00	1986.0
274.00	5380.0
275.00	29808.0
276.00	3838.0
277.00	2803.0
278.00	478.0
281.10	348.0
282.10	202.0
283.00	357.0
283.90	331.0
285.00	578.0
288.90	169.0
293.00	664.0
296.00	8833.0
297.00	1374.0
303.00	927.0
303.90	198.0
308.00	205.0
313.90	332.0
314.90	1084.0
316.10	486.0
321.00	294.0
323.00	2855.0
324.00	446.0
326.90	431.0
327.90	198.0
333.00	222.0
334.00	1873.0
334.90	458.0
341.00	368.0
345.90	510.0
352.00	620.0
352.90	575.0
354.00	735.0
355.00	333.0
365.00	3442.0
365.90	502.0
371.00	250.0
371.90	1378.0
373.00	343.0
383.00	341.0
383.80	157.0
390.00	210.0
401.90	514.0
403.10	628.0
404.00	368.0
420.80	484.0
422.00	659.0
423.00	4009.0
423.90	709.0
441.00	11762.0
442.00	71864.0
443.00	13821.0
443.90	1114.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
37.10	264.0
37.90	668.0
39.00	4682.0
40.00	2277.0
41.10	446.0
42.90	339.0
44.00	3165.0
44.90	292.0
50.00	15013.0
51.00	57752.0
52.00	3145.0
55.00	449.0
56.00	1898.0
57.00	4260.0
58.00	218.0
61.00	903.0
62.00	758.0
63.00	2346.0
64.00	438.0
65.00	1166.0
67.00	187.0
69.00	59864.0
70.00	372.0
71.10	220.0
73.00	1541.0
74.00	6137.0
75.00	8998.0
76.10	3214.0
77.00	64184.0
78.00	4476.0
79.00	4826.0
79.90	3504.0
81.00	5061.0
82.00	1103.0
83.00	1232.0
83.90	243.0
84.90	954.0
86.00	1638.0
87.00	630.0
87.90	210.0
90.90	1008.0
92.10	1226.0
93.00	8393.0
93.90	586.0
95.00	253.0
95.90	405.0
98.00	6551.0
99.00	4565.0
99.90	427.0
101.00	2371.0
103.00	945.0
103.90	1844.0
105.00	1607.0
107.00	18888.0
108.00	3242.0
110.00	35616.0
111.00	5511.0
112.00	709.0
112.80	206.0
115.20	161.0
116.00	1234.0
117.00	15923.0
118.00	1186.0
118.90	223.0
119.90	307.0
122.00	1481.0
123.00	1999.0
123.90	775.0
125.00	917.0
127.00	69248.0
128.00	5086.0
129.00	26296.0
129.90	1960.0
130.90	439.0
132.00	250.0
133.10	312.0
134.00	790.0
135.00	2361.0
136.00	1029.0
137.00	1180.0
138.00	313.0
138.90	258.0
139.90	412.0
141.00	3628.0
142.00	1051.0
143.00	868.0
143.80	195.0
146.00	724.0
147.00	2214.0
148.00	4762.0
149.00	755.0
150.00	220.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

151.00	525.0
151.90	201.0
152.90	1028.0
154.00	1022.0
155.00	2162.0
156.00	3097.0
157.00	635.0
158.00	786.0
159.00	589.0
160.00	1097.0
161.00	1605.0
161.90	473.0
163.00	296.0
163.90	178.0
165.00	1247.0
166.00	1225.0
167.00	7350.0
168.00	3278.0
169.00	748.0
170.00	294.0
170.70	378.0
172.00	602.0
172.90	927.0
174.00	1439.0
175.00	2882.0
176.10	730.0
176.80	1304.0
178.00	507.0
179.00	5505.0
180.00	3843.0
181.00	1541.0
182.00	264.0
183.00	261.0
183.90	421.0
185.00	2689.0
186.00	20688.0
187.00	5725.0
188.00	634.0
188.90	1187.0
189.90	254.0
191.00	866.0
191.90	1849.0
193.00	2341.0
194.00	442.0
195.00	314.0
196.00	4185.0
197.90	154880.0
199.00	10735.0
200.00	765.0
201.60	766.0
203.00	1133.0
204.00	5436.0
205.00	10030.0
206.00	38800.0
207.00	7282.0
208.00	1819.0
208.90	587.0
210.10	659.0
211.00	1409.0
212.00	301.0
215.00	448.0
216.10	911.0
217.00	11014.0
218.00	1144.0
218.90	162.0
221.00	6174.0
222.10	1394.0
223.00	2426.0
224.00	21976.0
225.00	5320.0
226.00	773.0
227.00	10811.0
228.00	1381.0
228.90	1949.0
229.90	293.0
230.90	879.0
232.10	171.0
233.00	261.0
233.90	450.0
235.00	541.0
236.00	491.0
236.90	704.0
239.00	492.0
240.00	232.0
240.90	487.0
242.00	1115.0
243.10	1400.0
244.00	17720.0
245.00	2361.0
246.00	3500.0
247.00	882.0
248.00	205.0
248.80	697.0
250.90	199.0
252.00	251.0
253.00	536.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

255.00	95232.0
256.00	13411.0
257.00	1227.0
258.00	4929.0
258.90	872.0
264.00	214.0
265.00	1977.0
266.00	318.0
271.00	181.0
272.00	267.0
273.00	2702.0
274.00	6949.0
275.00	42840.0
276.00	5440.0
277.00	3640.0
277.90	699.0
281.00	287.0
282.20	241.0
283.10	510.0
284.00	304.0
284.90	746.0
292.00	159.0
293.00	1066.0
296.00	11909.0
297.00	1716.0
297.90	181.0
302.00	257.0
303.00	1458.0
304.00	477.0
308.00	167.0
313.90	428.0
315.10	1195.0
316.00	559.0
320.90	332.0
322.10	205.0
323.00	4159.0
323.90	750.0
327.00	1023.0
327.90	381.0
331.90	252.0
332.90	301.0
334.00	2458.0
335.10	633.0
341.00	489.0
346.00	821.0
352.00	1267.0
353.00	749.0
354.00	1440.0
355.00	336.0
364.90	5113.0
365.90	742.0
370.90	292.0
372.00	2100.0
373.00	528.0
383.00	526.0
390.00	238.0
391.10	212.0
391.80	210.0
401.00	189.0
402.00	1017.0
403.00	1121.0
404.00	531.0
421.00	1065.0
422.00	1243.0
423.00	6339.0
423.90	977.0
425.00	223.0
441.00	17368.0
442.00	109368.0
443.00	21848.0
444.10	2049.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

m/z	Abundance
38.10	470.0
39.10	2493.0
40.00	2205.0
41.00	485.0
43.00	353.0
44.00	3044.0
45.00	196.0
50.00	9070.0
51.00	35528.0
52.00	1753.0
55.00	584.0
55.90	1076.0
57.00	2372.0
60.90	589.0
62.00	386.0
63.00	1326.0
63.90	153.0
65.00	634.0
67.00	188.0
69.00	35464.0
70.00	300.0
71.00	156.0
73.00	1042.0
74.00	3973.0
75.00	6345.0
76.10	2016.0
77.00	39248.0
78.00	2681.0
79.00	2959.0
80.00	2285.0
81.00	2795.0
82.00	634.0
83.00	694.0
84.00	182.0
85.00	375.0
86.00	857.0
87.00	387.0
91.00	716.0
92.00	834.0
93.00	5117.0
94.00	293.0
95.10	243.0
96.00	358.0
98.00	4080.0
99.00	2825.0
100.10	248.0
101.00	1350.0
102.90	576.0
104.00	1194.0
104.90	1145.0
106.00	427.0
107.00	12804.0
108.00	2071.0
110.00	23928.0
111.00	3622.0
111.90	400.0
113.00	274.0
116.00	758.0
117.00	10706.0
118.00	844.0
118.80	181.0
120.00	158.0
122.00	953.0
123.00	1460.0
124.00	587.0
125.00	586.0
127.00	44856.0
128.00	3502.0
129.00	16816.0
129.90	1502.0
131.00	451.0
132.60	183.0
132.90	182.0
133.90	507.0
135.00	1661.0
136.00	568.0
137.00	902.0
137.90	160.0
139.90	221.0
141.00	2378.0
141.90	776.0
142.90	436.0
143.80	193.0
146.00	380.0
147.00	1796.0
148.00	2624.0
148.90	579.0
150.00	247.0
151.10	344.0
151.60	178.0
152.20	155.0
152.90	778.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

154.00	822.0
154.90	1233.0
156.00	1784.0
157.10	461.0
158.00	441.0
159.00	328.0
160.00	862.0
161.00	1330.0
161.90	271.0
164.10	232.0
165.00	1114.0
166.00	903.0
167.00	4704.0
168.00	2682.0
168.90	489.0
170.10	185.0
171.00	237.0
172.00	609.0
173.00	642.0
174.00	1133.0
175.00	2099.0
176.00	504.0
176.90	1029.0
178.10	366.0
178.90	3566.0
180.00	2526.0
181.00	1480.0
183.90	322.0
185.00	1923.0
186.00	15598.0
187.00	4147.0
188.00	519.0
189.00	918.0
190.90	700.0
192.00	1201.0
193.00	1465.0
194.00	403.0
195.00	270.0
196.00	3093.0
198.00	111968.0
199.00	7699.0
200.00	588.0
201.40	585.0
202.90	865.0
204.00	3650.0
205.00	7273.0
206.00	29704.0
207.00	5380.0
208.00	1249.0
209.00	605.0
210.20	432.0
211.00	1129.0
214.90	285.0
215.80	530.0
217.00	8378.0
217.90	1075.0
221.00	5171.0
223.00	1762.0
224.00	16832.0
225.00	4316.0
226.00	507.0
227.00	6911.0
228.00	939.0
229.00	1576.0
229.80	166.0
231.00	652.0
232.80	192.0
234.10	425.0
234.90	516.0
235.80	453.0
237.00	517.0
238.00	152.0
239.00	277.0
239.90	229.0
241.00	420.0
242.00	965.0
243.10	966.0
244.00	13746.0
245.10	1791.0
246.00	2681.0
246.90	576.0
248.10	180.0
249.00	411.0
251.10	207.0
253.00	310.0
255.00	70008.0
256.00	9578.0
257.00	873.0
258.00	4655.0
258.90	726.0
260.00	161.0
265.00	1326.0
266.00	238.0
267.00	203.0
271.00	200.0
273.00	2354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253

274.00	5464.0
275.00	32512.0
276.00	4243.0
277.00	2566.0
277.90	472.0
281.00	493.0
282.00	260.0
283.00	458.0
284.00	251.0
285.00	586.0
290.80	183.0
293.00	572.0
295.00	213.0
296.00	8865.0
297.00	1489.0
302.00	175.0
303.00	1183.0
304.10	266.0
308.90	165.0
314.00	525.0
314.90	1133.0
316.00	657.0
317.10	166.0
320.90	239.0
323.10	3063.0
324.00	536.0
327.00	592.0
328.00	297.0
332.00	232.0
332.70	186.0
334.00	2046.0
335.00	649.0
340.90	346.0
342.00	226.0
346.00	780.0
352.00	1084.0
353.00	648.0
354.10	1060.0
355.10	355.0
364.90	4701.0
365.90	938.0
370.00	152.0
371.00	359.0
372.10	1814.0
373.00	378.0
383.00	649.0
383.90	152.0
389.90	202.0
391.00	281.0
402.00	713.0
402.90	915.0
403.90	307.0
421.00	897.0
421.90	802.0
423.00	5563.0
423.90	1177.0
439.80	168.0
441.00	14733.0
442.00	98064.0
443.00	20056.0
443.90	1675.0

Instrument :
BNA_N
ClientSampleId :
DFTPP253