

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR021419\
 Data File : VR026317.D
 Acq On : 14 Feb 2019 17:57
 Operator : SY/MD
 Sample : K1488-07
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ESXD3

Quant Time: Feb 15 04:36:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 14 14:16:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	545809	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	469063	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	130050	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	229521	4.61	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.20%
7) Chloroethane-d5	2.63	69	195260	4.79	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.80%
11) 1,1-Dichloroethene-d2	3.61	63	431820	3.61	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.20%
20) 2-Butanone-d5	6.59	46	176212	55.53	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	111.06%
24) Chloroform-d	7.20	84	387466	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.80%
26) 1,2-Dichloroethane-d4	7.89	65	153863	5.11	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.20%
32) Benzene-d6	7.85	84	763696	4.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.40%
36) 1,2-Dichloropropane-d6	8.89	67	196524	4.79	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.80%
41) Toluene-d8	9.97	98	710246	4.75	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.00%
43) trans-1,3-Dichloropropene-	10.24	79	48644	4.40	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.00%
46) 2-Hexanone-d5	10.59	63	148132	59.05	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	118.10%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	74964	4.81	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	96.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	135345	6.22	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	124.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR021419\
Data File : VR026317.D
Acq On : 14 Feb 2019 17:57
Operator : SY/MD
Sample : K1488-07
Misc : 25mL/MSVOA R/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ESXD3

Quant Time: Feb 15 04:36:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 14 14:16:44 2019
Response via : Initial Calibration

