

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR022019\
 Data File : VR026345.D
 Acq On : 20 Feb 2019 10:12
 Operator : SY/MD
 Sample : VR0220WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK83

Quant Time: Feb 21 07:46:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 21 07:38:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	601897	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	483886	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	140542	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.16	65	221343	4.65	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.00%
7) Chloroethane-d5	2.63	69	199067	4.79	ug/L	0.00
Spiked Amount	7.000	Range	65 - 130	Recovery	=	68.43%
11) 1,1-Dichloroethene-d2	3.61	63	419782	3.50	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.00%
20) 2-Butanone-d5	6.59	46	167787	46.38	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.76%
24) Chloroform-d	7.20	84	393002	4.53	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	7.90	65	147751	4.58	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.60%
32) Benzene-d6	7.85	84	752876	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.80%
36) 1,2-Dichloropropane-d6	8.89	67	191240	4.90	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.00%
41) Toluene-d8	9.97	98	672971	4.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.60%
43) trans-1,3-Dichloropropene-	10.23	79	44875	4.25	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	85.00%
46) 2-Hexanone-d5	10.59	63	128566	48.58	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.16%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	67719	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	124248	5.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR022019\
Data File : VR026345.D
Acq On : 20 Feb 2019 10:12
Operator : SY/MD
Sample : VR0220WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK83

Quant Time: Feb 21 07:46:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 21 07:38:25 2019
Response via : Initial Calibration

