

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR022019\
 Data File : VR026346.D
 Acq On : 20 Feb 2019 11:05
 Operator : SY/MD
 Sample : K1488-01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ESXC7

Quant Time: Feb 21 07:48:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Feb 21 07:38:25 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	524149	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	432797	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	119458	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	215829	5.21	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	104.20%
7) Chloroethane-d5	2.63	69	197691	5.47	ug/L	0.00
Spiked Amount	7.000	Range	65 - 130	Recovery	=	78.14%
11) 1,1-Dichloroethene-d2	3.60	63	417679	4.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.00%
20) 2-Butanone-d5	6.58	46	155975	49.51	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.02%
24) Chloroform-d	7.20	84	375980	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.60%
26) 1,2-Dichloroethane-d4	7.89	65	144188	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	7.85	84	748682	5.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	109.80%
36) 1,2-Dichloropropane-d6	8.89	67	182509	5.23	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.60%
41) Toluene-d8	9.96	98	686610	5.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.20%
43) trans-1,3-Dichloropropene-	10.23	79	42059	4.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.00%
46) 2-Hexanone-d5	10.59	63	120907	51.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.16%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65516	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	121821	6.55	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	131.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR022019\
Data File : VR026346.D
Acq On : 20 Feb 2019 11:05
Operator : SY/MD
Sample : K1488-01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ESXC7

Quant Time: Feb 21 07:48:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Feb 21 07:38:25 2019
Response via : Initial Calibration

