

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR022119\
 Data File : VR026359.D
 Acq On : 21 Feb 2019 13:51
 Operator : SY/MD
 Sample : K1531-04
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0TR9

Quant Time: Feb 22 06:34:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Feb 22 06:18:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	510032	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	435190	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	109488	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	168758	4.18	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.60%
7) Chloroethane-d5	2.62	69	169001	4.80	ug/L	0.00
Spiked Amount	7.000	Range	65 - 130	Recovery	=	68.57%
11) 1,1-Dichloroethene-d2	3.60	63	339437	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	6.59	46	157290	51.31	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.62%
24) Chloroform-d	7.20	84	340812	4.64	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.80%
26) 1,2-Dichloroethane-d4	7.89	65	133351	4.88	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	97.60%
32) Benzene-d6	7.85	84	620556	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.40%
36) 1,2-Dichloropropane-d6	8.89	67	167311	4.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.40%
41) Toluene-d8	9.97	98	544381	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.00%
43) trans-1,3-Dichloropropene-	10.23	79	38727	4.07	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.40%
46) 2-Hexanone-d5	10.59	63	121753	51.15	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.30%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64792	4.62	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	108097	6.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	126.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR022119\
Data File : VR026359.D
Acq On : 21 Feb 2019 13:51
Operator : SY/MD
Sample : K1531-04
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0TR9

Quant Time: Feb 22 06:34:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR021919WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Feb 22 06:18:00 2019
Response via : Initial Calibration

