

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040419\
 Data File : VR026445.D
 Acq On : 4 Apr 2019 14:15
 Operator : SY/MD
 Sample : K2260-03
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ETJ51

Quant Time: Apr 05 07:38:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 05 07:16:54 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	465539	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	388394	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	112379	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	192625	5.45	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.00%
7) Chloroethane-d5	2.62	69	162380	5.51	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.20%
11) 1,1-Dichloroethene-d2	3.60	63	380892	4.16	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.20%
20) 2-Butanone-d5	6.59	46	154933	49.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.26%
24) Chloroform-d	7.20	84	335418	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	7.89	65	143230	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.40%
32) Benzene-d6	7.85	84	624847	5.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.60%
36) 1,2-Dichloropropane-d6	8.89	67	160573	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	9.97	98	566940	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.00%
43) trans-1,3-Dichloropropene-	10.23	79	37602	4.41	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.20%
46) 2-Hexanone-d5	10.59	63	126508	55.33	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	110.66%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	58901	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	102369	6.02	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040419\
Data File : VR026445.D
Acq On : 4 Apr 2019 14:15
Operator : SY/MD
Sample : K2260-03
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ETJ51

Quant Time: Apr 05 07:38:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 05 07:16:54 2019
Response via : Initial Calibration

