

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040519\
 Data File : VR026456.D
 Acq On : 5 Apr 2019 11:03
 Operator : SY/MD
 Sample : K2271-05
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 DB7C8

Quant Time: Apr 06 06:26:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 06 05:54:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	513599	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	421825	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	122608	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	145334	3.72	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.40%
7) Chloroethane-d5	2.63	69	129169	3.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.40%
11) 1,1-Dichloroethene-d2	3.61	63	289540	2.87	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.40%#
20) 2-Butanone-d5	6.59	46	125712	36.13	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	72.26%
24) Chloroform-d	7.20	84	274755	3.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	74.00%
26) 1,2-Dichloroethane-d4	7.89	65	118158	3.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.80%
32) Benzene-d6	7.85	84	488537	3.66	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	73.20%
36) 1,2-Dichloropropane-d6	8.89	67	131172	3.76	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	75.20%
41) Toluene-d8	9.97	98	426958	3.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.80%
43) trans-1,3-Dichloropropene-	10.23	79	28805	3.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.20%
46) 2-Hexanone-d5	10.59	63	93832	37.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.56%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	47734	3.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	70.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	82005	4.42	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040519\
Data File : VR026456.D
Acq On : 5 Apr 2019 11:03
Operator : SY/MD
Sample : K2271-05
Misc : 25mL/MSVOA R/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
DB7C8

Quant Time: Apr 06 06:26:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 06 05:54:53 2019
Response via : Initial Calibration

