

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040519\
 Data File : VR026461.D
 Acq On : 5 Apr 2019 13:44
 Operator : SY/MD
 Sample : K2271-17
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 DB7D8

Quant Time: Apr 06 06:41:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 06 05:54:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	481684	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	395633	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	112700	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	160300	4.38	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.60%
7) Chloroethane-d5	2.62	69	140755	4.61	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.20%
11) 1,1-Dichloroethene-d2	3.60	63	312899	3.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.00%
20) 2-Butanone-d5	6.58	46	155121	47.54	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.08%
24) Chloroform-d	7.20	84	301106	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
26) 1,2-Dichloroethane-d4	7.89	65	126975	4.34	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.80%
32) Benzene-d6	7.85	84	541757	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
36) 1,2-Dichloropropane-d6	8.89	67	149442	4.56	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.20%
41) Toluene-d8	9.97	98	476054	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.00%
43) trans-1,3-Dichloropropene-	10.23	79	29076	3.35	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.00%
46) 2-Hexanone-d5	10.59	63	112861	48.46	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	96.92%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	55734	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	89612	5.25	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040519\
Data File : VR026461.D
Acq On : 5 Apr 2019 13:44
Operator : SY/MD
Sample : K2271-17
Misc : 25mL/MSVOA R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
DB7D8

Quant Time: Apr 06 06:41:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 06 05:54:53 2019
Response via : Initial Calibration

