

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040519\
 Data File : VR026468.D
 Acq On : 5 Apr 2019 17:59
 Operator : SY/MD
 Sample : K2271-03
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 DB7C6

Quant Time: Apr 07 13:27:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 06 05:54:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	557137	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	448912	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	127641	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.13	65	163111	3.85	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.00%
7) Chloroethane-d5	2.62	69	139932	3.96	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.20%
11) 1,1-Dichloroethene-d2	3.60	63	325807	2.98	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.60%#
20) 2-Butanone-d5	6.58	46	158471	41.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	83.98%
24) Chloroform-d	7.19	84	306356	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.20%
26) 1,2-Dichloroethane-d4	7.89	65	129900	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
32) Benzene-d6	7.85	84	549654	3.87	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.40%
36) 1,2-Dichloropropane-d6	8.89	67	148936	4.01	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	80.20%
41) Toluene-d8	9.97	98	489344	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
43) trans-1,3-Dichloropropene-	10.23	79	31139	3.16	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	63.20%
46) 2-Hexanone-d5	10.59	63	116921	44.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.48%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	53905	3.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	95052	4.92	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040519\
Data File : VR026468.D
Acq On : 5 Apr 2019 17:59
Operator : SY/MD
Sample : K2271-03
Misc : 25mL/MSVOA R/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
DB7C6

Quant Time: Apr 07 13:27:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 06 05:54:53 2019
Response via : Initial Calibration

