

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040619\
 Data File : VR026478.D
 Acq On : 6 Apr 2019 16:35
 Operator : SY/MD
 Sample : K2260-07
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 ETJ52

Quant Time: Apr 06 23:47:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 06 23:34:06 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	539418	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	442751	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	131330	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	164870	4.02	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	80.40%
7) Chloroethane-d5	2.62	69	145443	4.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	85.20%
11) 1,1-Dichloroethene-d2	3.60	63	324235	3.06	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.20%
20) 2-Butanone-d5	6.58	46	168935	46.23	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.46%
24) Chloroform-d	7.20	84	335295	4.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.00%
26) 1,2-Dichloroethane-d4	7.90	65	139615	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.40%
32) Benzene-d6	7.85	84	598669	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
36) 1,2-Dichloropropane-d6	8.89	67	163664	4.47	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.40%
41) Toluene-d8	9.97	98	516429	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
43) trans-1,3-Dichloropropene-	10.23	79	29264	3.01	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	60.20%
46) 2-Hexanone-d5	10.59	63	128306	49.22	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.44%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	61296	4.29	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.80%
64) 1,2-Dichlorobenzene-d4	13.51	152	98390	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040619\
Data File : VR026478.D
Acq On : 6 Apr 2019 16:35
Operator : SY/MD
Sample : K2260-07
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
ETJ52

Quant Time: Apr 06 23:47:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 06 23:34:06 2019
Response via : Initial Calibration

