

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026494.D
 Acq On : 7 Apr 2019 13:51
 Operator : SY/MD
 Sample : K2282-10
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZ08

Quant Time: Apr 08 04:07:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	515263	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	424100	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	117417	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	143745	3.67	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.40%
7) Chloroethane-d5	2.63	69	130949	4.01	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.20%
11) 1,1-Dichloroethene-d2	3.60	63	307173	3.03	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.60%
20) 2-Butanone-d5	6.59	46	154524	44.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.54%
24) Chloroform-d	7.20	84	317445	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
26) 1,2-Dichloroethane-d4	7.89	65	133031	4.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.00%
32) Benzene-d6	7.85	84	562825	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.80%
36) 1,2-Dichloropropane-d6	8.89	67	154245	4.40	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.00%
41) Toluene-d8	9.97	98	475488	3.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.40%
43) trans-1,3-Dichloropropene-	10.23	79	26590	2.86	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	57.20%
46) 2-Hexanone-d5	10.59	63	122696	49.14	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.28%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	58924	4.30	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	94565	5.32	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026494.D
Acq On : 7 Apr 2019 13:51
Operator : SY/MD
Sample : K2282-10
Misc : 25mL/MSVOA R/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZ08

Quant Time: Apr 08 04:07:14 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

