

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026500.D
 Acq On : 7 Apr 2019 17:04
 Operator : SY/MD
 Sample : K2282-16
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZ14

Quant Time: Apr 08 04:29:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	391430	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	329059	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	93450	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	178340	6.00	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.00%
7) Chloroethane-d5	2.63	69	150579	6.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	121.40%
11) 1,1-Dichloroethene-d2	3.60	63	353856	4.60	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.00%
20) 2-Butanone-d5	6.59	46	168661	63.61	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	127.22%
24) Chloroform-d	7.20	84	335297	5.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	118.60%
26) 1,2-Dichloroethane-d4	7.89	65	149051	6.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	125.60%
32) Benzene-d6	7.85	84	620406	5.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	119.20%
36) 1,2-Dichloropropane-d6	8.89	67	165841	6.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	121.80%
41) Toluene-d8	9.97	98	548646	5.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	116.60%
43) trans-1,3-Dichloropropene-	10.23	79	30095	4.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.40%
46) 2-Hexanone-d5	10.59	63	131668	67.97	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	135.94%#
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60609	5.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	114.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	103867	7.34	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	146.80%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026500.D
Acq On : 7 Apr 2019 17:04
Operator : SY/MD
Sample : K2282-16
Misc : 25mL/MSVOA R/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZ14

Quant Time: Apr 08 04:29:30 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

