

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026490.D
 Acq On : 7 Apr 2019 11:43
 Operator : SY/MD
 Sample : K2271-25
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK02

Quant Time: Apr 08 03:45:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	450875	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	362404	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	106185	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	182191	5.32	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.40%
7) Chloroethane-d5	2.62	69	158698	5.56	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	111.20%
11) 1,1-Dichloroethene-d2	3.60	63	362626	4.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.80%
20) 2-Butanone-d5	6.59	46	157992	51.73	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.46%
24) Chloroform-d	7.20	84	344227	5.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.80%
26) 1,2-Dichloroethane-d4	7.89	65	143528	5.25	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.00%
32) Benzene-d6	7.85	84	621819	5.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.40%
36) 1,2-Dichloropropane-d6	8.90	67	166992	5.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	111.40%
41) Toluene-d8	9.97	98	565257	5.45	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	109.00%
43) trans-1,3-Dichloropropene-	10.23	79	34417	4.33	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.60%
46) 2-Hexanone-d5	10.59	63	128525	60.24	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	120.48%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	59413	5.08	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	104458	6.50	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	130.00%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026490.D
Acq On : 7 Apr 2019 11:43
Operator : SY/MD
Sample : K2271-25
Misc : 25mL/MSVOA R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK02

Quant Time: Apr 08 03:45:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

