

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026492.D
 Acq On : 7 Apr 2019 12:47
 Operator : SY/MD
 Sample : K2282-08
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZ06

Quant Time: Apr 08 09:31:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	480924	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	399178	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	112900	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	156430	4.28	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.60%
7) Chloroethane-d5	2.63	69	137776	4.52	ug/L	0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	3.60	63	315460	3.34	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.80%
20) 2-Butanone-d5	6.59	46	153649	47.16	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.32%
24) Chloroform-d	7.20	84	318737	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	7.89	65	136694	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	7.85	84	555511	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.00%
36) 1,2-Dichloropropane-d6	8.89	67	158816	4.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.20%
41) Toluene-d8	9.97	98	480986	4.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.20%
43) trans-1,3-Dichloropropene-	10.23	79	26015	2.97	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	59.40%
46) 2-Hexanone-d5	10.59	63	120282	51.18	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	102.36%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	57375	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	95491	5.59	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026492.D
Acq On : 7 Apr 2019 12:47
Operator : SY/MD
Sample : K2282-08
Misc : 25mL/MSVOA R/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZ06

Quant Time: Apr 08 09:31:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

