

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026497.D
 Acq On : 7 Apr 2019 15:27
 Operator : SY/MD
 Sample : K2282-13
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZ11

Quant Time: Apr 08 09:33:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	462736	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	387376	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	110761	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	143114	4.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.40%
7) Chloroethane-d5	2.63	69	136438	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.00%
11) 1,1-Dichloroethene-d2	3.61	63	296111	3.26	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.20%
20) 2-Butanone-d5	6.59	46	154916	49.42	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.84%
24) Chloroform-d	7.20	84	306678	4.59	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.80%
26) 1,2-Dichloroethane-d4	7.90	65	132083	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
32) Benzene-d6	7.85	84	530743	4.33	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.60%
36) 1,2-Dichloropropane-d6	8.90	67	147178	4.59	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	91.80%
41) Toluene-d8	9.97	98	457117	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	10.24	79	24802	2.92	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	58.40%
46) 2-Hexanone-d5	10.59	63	122882	53.88	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.76%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	59614	4.77	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	92578	5.52	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026497.D
Acq On : 7 Apr 2019 15:27
Operator : SY/MD
Sample : K2282-13
Misc : 25mL/MSVOA R/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZ11

Quant Time: Apr 08 09:33:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

