

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040719\
 Data File : VR026498.D
 Acq On : 7 Apr 2019 15:59
 Operator : SY/MD
 Sample : K2282-14
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZ12

Quant Time: Apr 08 09:33:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Apr 08 02:57:29 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	478059	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	408713	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	115747	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	148625	4.09	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.80%
7) Chloroethane-d5	2.63	69	136446	4.50	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.00%
11) 1,1-Dichloroethene-d2	3.60	63	305183	3.25	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	6.59	46	159951	49.39	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.78%
24) Chloroform-d	7.20	84	319522	4.63	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.60%
26) 1,2-Dichloroethane-d4	7.89	65	135884	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.60%
32) Benzene-d6	7.85	84	548733	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
36) 1,2-Dichloropropane-d6	8.89	67	152897	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	90.40%
41) Toluene-d8	9.97	98	469176	4.01	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.20%
43) trans-1,3-Dichloropropene-	10.24	79	24171	2.69	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	53.80%#
46) 2-Hexanone-d5	10.59	63	127753	53.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	106.18%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60258	4.57	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.40%
64) 1,2-Dichlorobenzene-d4	13.51	152	97667	5.57	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	111.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040719\
Data File : VR026498.D
Acq On : 7 Apr 2019 15:59
Operator : SY/MD
Sample : K2282-14
Misc : 25mL/MSVOA R/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZ12

Quant Time: Apr 08 09:33:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Apr 08 02:57:29 2019
Response via : Initial Calibration

