

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040819\  
 Data File : VR026512.D  
 Acq On : 8 Apr 2019 17:20  
 Operator : SY/MD  
 Sample : K2271-25  
 Misc : 25mL/MSVOA R/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_R  
 ClientSampleId :  
 VHBLK02

Quant Time: Apr 09 02:03:23 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_R\METHODS\SOMRTR040319WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Tue Apr 09 01:39:02 2019  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 8.46  | 114  | 661330   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 11.29 | 117  | 555374   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 13.23 | 152  | 169700   | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 2.14   | 65    | 199539   | 3.97     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 79.40%  |
| 7) Chloroethane-d5             | 2.63   | 69    | 172912   | 4.13     | ug/L | 0.01    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 82.60%  |
| 11) 1,1-Dichloroethene-d2      | 3.63   | 63    | 408958   | 3.15     | ug/L | 0.02    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 63.00%  |
| 20) 2-Butanone-d5              | 6.58   | 46    | 228379   | 50.98    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 101.96% |
| 24) Chloroform-d               | 7.20   | 84    | 423533   | 4.43     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 88.60%  |
| 26) 1,2-Dichloroethane-d4      | 7.90   | 65    | 190854   | 4.76     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 95.20%  |
| 32) Benzene-d6                 | 7.85   | 84    | 773150   | 4.40     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 88.00%  |
| 36) 1,2-Dichloropropane-d6     | 8.89   | 67    | 211687   | 4.61     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 92.20%  |
| 41) Toluene-d8                 | 9.97   | 98    | 714454   | 4.49     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 89.80%  |
| 43) trans-1,3-Dichloropropene- | 10.23  | 79    | 47040    | 3.86     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 77.20%  |
| 46) 2-Hexanone-d5              | 10.59  | 63    | 190061   | 58.13    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 116.26% |
| 57) 1,1,2,2-Tetrachloroethane- | 12.36  | 84    | 83475    | 4.65     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 93.00%  |
| 64) 1,2-Dichlorobenzene-d4     | 13.52  | 152   | 141042   | 5.49     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 109.80% |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040819\  
Data File : VR026512.D  
Acq On : 8 Apr 2019 17:20  
Operator : SY/MD  
Sample : K2271-25  
Misc : 25mL/MSVOA R/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_R  
ClientSampleId :  
VHBLK02

Quant Time: Apr 09 02:03:23 2019  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_R\METHODS\SOMRTR040319WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Tue Apr 09 01:39:02 2019  
Response via : Initial Calibration

