

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026523.D
 Acq On : 9 Apr 2019 19:18
 Operator : SY/MD
 Sample : K2285-07
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZC5

Quant Time: Apr 10 06:53:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 05:42:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	645920	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	518655	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	154965	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	171340	3.49	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	69.80%
7) Chloroethane-d5	2.63	69	163196	3.99	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	79.80%
11) 1,1-Dichloroethene-d2	3.60	63	357248	2.81	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.20%#
20) 2-Butanone-d5	6.59	46	177446	40.55	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.10%
24) Chloroform-d	7.20	84	356665	3.82	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	76.40%
26) 1,2-Dichloroethane-d4	7.90	65	158225	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.80%
32) Benzene-d6	7.85	84	647907	3.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	79.00%
36) 1,2-Dichloropropane-d6	8.89	67	175012	4.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.60%
41) Toluene-d8	9.97	98	566329	3.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.20%
43) trans-1,3-Dichloropropene-	10.24	79	32433	2.85	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	57.00%
46) 2-Hexanone-d5	10.59	63	139223	45.60	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.20%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	66673	3.98	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	115299	4.91	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026523.D
Acq On : 9 Apr 2019 19:18
Operator : SY/MD
Sample : K2285-07
Misc : 25mL/MSVOA R/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZC5

Quant Time: Apr 10 06:53:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 05:42:05 2019
Response via : Initial Calibration

