

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026538.D
 Acq On : 10 Apr 2019 4:22
 Operator : SY/MD
 Sample : K2285-20
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZD8

Quant Time: Apr 10 08:03:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 07:28:30 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	513880	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	427695	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	121798	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	176400	4.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.40%
7) Chloroethane-d5	2.63	69	151563	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.20%
11) 1,1-Dichloroethene-d2	3.61	63	361172	3.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.60%
20) 2-Butanone-d5	6.59	46	184902	53.12	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.24%
24) Chloroform-d	7.20	84	364516	4.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	7.89	65	159583	5.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.40%
32) Benzene-d6	7.85	84	666605	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.60%
36) 1,2-Dichloropropane-d6	8.90	67	179679	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	9.97	98	601805	4.92	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.40%
43) trans-1,3-Dichloropropene-	10.23	79	34707	3.70	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.00%
46) 2-Hexanone-d5	10.59	63	152616	60.61	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	121.22%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	72697	5.26	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	123042	6.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	133.40%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026538.D
Acq On : 10 Apr 2019 4:22
Operator : SY/MD
Sample : K2285-20
Misc : 25mL/MSVOA R/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZD8

Quant Time: Apr 10 08:03:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 07:28:30 2019
Response via : Initial Calibration

