

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026550.D
 Acq On : 10 Apr 2019 10:45
 Operator : SY/MD
 Sample : K2285-04DL 4X
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZC2DL

Quant Time: Apr 11 04:15:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 11 03:05:08 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	502790	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	420031	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	117196	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	167524	4.39	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.80%
7) Chloroethane-d5	2.62	69	151768	4.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	3.60	63	361360	3.66	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	73.20%
20) 2-Butanone-d5	6.58	46	168284	49.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.82%
24) Chloroform-d	7.20	84	347668	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.80%
26) 1,2-Dichloroethane-d4	7.89	65	151858	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
32) Benzene-d6	7.85	84	621344	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
36) 1,2-Dichloropropane-d6	8.89	67	169942	4.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.80%
41) Toluene-d8	9.97	98	572535	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
43) trans-1,3-Dichloropropene-	10.23	79	36685	3.98	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.60%
46) 2-Hexanone-d5	10.59	63	135590	54.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.66%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	65990	4.87	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.40%
64) 1,2-Dichlorobenzene-d4	13.52	152	114104	6.43	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	128.60%#

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026550.D
Acq On : 10 Apr 2019 10:45
Operator : SY/MD
Sample : K2285-04DL 4X
Misc : 25mL/MSVOA R/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZC2DL

Quant Time: Apr 11 04:15:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 11 03:05:08 2019
Response via : Initial Calibration

