

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026553.D
 Acq On : 10 Apr 2019 15:32
 Operator : SY/MD
 Sample : VR0409WBL03
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK96

Quant Time: Apr 11 04:27:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 11 04:17:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	553874	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	457188	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	134138	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	170799	4.06	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.20%
7) Chloroethane-d5	2.62	69	153163	4.36	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.20%
11) 1,1-Dichloroethene-d2	3.60	63	354297	3.25	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.00%
20) 2-Butanone-d5	6.59	46	151791	40.46	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	80.92%
24) Chloroform-d	7.20	84	337496	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
26) 1,2-Dichloroethane-d4	7.89	65	142097	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.60%
32) Benzene-d6	7.85	84	611497	4.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.60%
36) 1,2-Dichloropropane-d6	8.89	67	162556	4.30	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.00%
41) Toluene-d8	9.97	98	545675	4.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.40%
43) trans-1,3-Dichloropropene-	10.23	79	34627	3.45	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.00%
46) 2-Hexanone-d5	10.59	63	122720	45.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.18%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	60639	4.11	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.20%
64) 1,2-Dichlorobenzene-d4	13.52	152	109557	5.39	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026553.D
Acq On : 10 Apr 2019 15:32
Operator : SY/MD
Sample : VR0409WBL03
Misc : 25mL/MSVOA R/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK96

Quant Time: Apr 11 04:27:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 11 04:17:38 2019
Response via : Initial Calibration

