

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026563.D
 Acq On : 10 Apr 2019 21:00
 Operator : SY/MD
 Sample : K2289-18
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZJ8

Quant Time: Apr 11 05:02:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 11 04:17:38 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	563563	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	464668	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	136811	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	150225	3.51	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	70.20%
7) Chloroethane-d5	2.63	69	139421	3.90	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.00%
11) 1,1-Dichloroethene-d2	3.61	63	306752	2.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.40%#
20) 2-Butanone-d5	6.59	46	155642	40.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	81.54%
24) Chloroform-d	7.20	84	329554	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.00%
26) 1,2-Dichloroethane-d4	7.89	65	139387	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.60%
32) Benzene-d6	7.85	84	571720	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.80%
36) 1,2-Dichloropropane-d6	8.90	67	156744	4.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.60%
41) Toluene-d8	9.97	98	503207	3.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.60%
43) trans-1,3-Dichloropropene-	10.23	79	32041	3.14	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.80%
46) 2-Hexanone-d5	10.59	63	125373	45.83	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.66%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	61310	4.09	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	111094	5.36	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026563.D
Acq On : 10 Apr 2019 21:00
Operator : SY/MD
Sample : K2289-18
Misc : 25mL/MSVOA R/WATER
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZJ8

Quant Time: Apr 11 05:02:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 11 04:17:38 2019
Response via : Initial Calibration

