

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026516.D
 Acq On : 9 Apr 2019 15:31
 Operator : SY/MD
 Sample : VR0409WBL01
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VBLK94

Quant Time: Apr 10 06:36:42 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 05:42:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	594840	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	487770	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	144558	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	195168	4.32	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	86.40%
7) Chloroethane-d5	2.62	69	174834	4.64	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.80%
11) 1,1-Dichloroethene-d2	3.61	63	389690	3.33	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.60%
20) 2-Butanone-d5	6.59	46	176386	43.77	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.54%
24) Chloroform-d	7.20	84	378891	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.20%
26) 1,2-Dichloroethane-d4	7.90	65	161128	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	7.85	84	690684	4.48	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.60%
36) 1,2-Dichloropropane-d6	8.89	67	186657	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	9.97	98	642517	4.60	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.00%
43) trans-1,3-Dichloropropene-	10.24	79	40724	3.80	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.00%
46) 2-Hexanone-d5	10.59	63	154348	53.75	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.50%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	71573	4.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.80%
64) 1,2-Dichlorobenzene-d4	13.52	152	129114	5.90	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	118.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026516.D
Acq On : 9 Apr 2019 15:31
Operator : SY/MD
Sample : VR0409WBL01
Misc : 25mL/MSVOA R/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VBLK94

Quant Time: Apr 10 06:36:42 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 05:42:05 2019
Response via : Initial Calibration

