

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026518.D
 Acq On : 9 Apr 2019 16:39
 Operator : SY/MD
 Sample : K2285-02
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZC0

Quant Time: Apr 10 06:40:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 05:42:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	661082	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	541675	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	161835	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	184796	3.68	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.60%
7) Chloroethane-d5	2.62	69	164637	3.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.60%
11) 1,1-Dichloroethene-d2	3.60	63	375093	2.89	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.80%#
20) 2-Butanone-d5	6.59	46	174352	38.93	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.86%
24) Chloroform-d	7.20	84	360489	3.78	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	75.60%
26) 1,2-Dichloroethane-d4	7.90	65	154230	3.84	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	76.80%
32) Benzene-d6	7.85	84	669116	3.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.00%
36) 1,2-Dichloropropane-d6	8.89	67	175684	3.92	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.40%
41) Toluene-d8	9.97	98	603516	3.89	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	77.80%
43) trans-1,3-Dichloropropene-	10.23	79	33477	2.82	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	56.40%
46) 2-Hexanone-d5	10.59	63	142798	44.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.56%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	68689	3.93	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	78.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	118490	4.84	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026518.D
Acq On : 9 Apr 2019 16:39
Operator : SY/MD
Sample : K2285-02
Misc : 25mL/MSVOA R/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZC0

Quant Time: Apr 10 06:40:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 05:42:05 2019
Response via : Initial Calibration

