

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
 Data File : VR026519.D
 Acq On : 9 Apr 2019 17:11
 Operator : SY/MD
 Sample : K2285-03
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 BEZC1

Quant Time: Apr 10 06:42:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 10 05:42:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	571612	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.28	117	472144	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.23	152	136619	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	167603	3.86	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.20%
7) Chloroethane-d5	2.62	69	160272	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.60%
11) 1,1-Dichloroethene-d2	3.61	63	346575	3.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	61.60%
20) 2-Butanone-d5	6.59	46	171536	44.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.60%
24) Chloroform-d	7.20	84	354055	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	7.89	65	152635	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	7.85	84	627132	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	8.89	67	170319	4.36	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.20%
41) Toluene-d8	9.97	98	547023	4.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.00%
43) trans-1,3-Dichloropropene-	10.23	79	31652	3.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	61.00%
46) 2-Hexanone-d5	10.59	63	135061	48.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.18%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	64446	4.23	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	84.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	111954	5.41	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR040919\
Data File : VR026519.D
Acq On : 9 Apr 2019 17:11
Operator : SY/MD
Sample : K2285-03
Misc : 25mL/MSVOA R/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
BEZC1

Quant Time: Apr 10 06:42:19 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR040319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 10 05:42:05 2019
Response via : Initial Calibration

