

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR041119\
 Data File : VR026578.D
 Acq On : 11 Apr 2019 20:49
 Operator : SY/MD
 Sample : K2289-21
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Apr 12 02:32:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR041119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 12 02:06:27 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	587388	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	477498	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	130675	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	179238	4.07	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	81.40%
7) Chloroethane-d5	2.62	69	166052	4.34	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	86.80%
11) 1,1-Dichloroethene-d2	3.60	63	375810	3.04	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.80%
20) 2-Butanone-d5	6.58	46	172557	41.20	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.40%
24) Chloroform-d	7.20	84	370338	4.08	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.60%
26) 1,2-Dichloroethane-d4	7.89	65	156697	4.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.00%
32) Benzene-d6	7.85	84	649224	4.20	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.00%
36) 1,2-Dichloropropane-d6	8.89	67	173753	4.18	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	83.60%
41) Toluene-d8	9.97	98	583775	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.40%
43) trans-1,3-Dichloropropene-	10.23	79	34238	3.28	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.60%
46) 2-Hexanone-d5	10.59	63	136869	43.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	87.70%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	63231	3.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	72.60%
64) 1,2-Dichlorobenzene-d4	13.52	152	111389	5.30	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR041119\
Data File : VR026578.D
Acq On : 11 Apr 2019 20:49
Operator : SY/MD
Sample : K2289-21
Misc : 25mL/MSVOA R/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Apr 12 02:32:45 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR041119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 12 02:06:27 2019
Response via : Initial Calibration

