

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR061618\
 Data File : VR025352.D
 Acq On : 17 Jun 2018 8:18
 Operator : SY/MD
 Sample : J3513-12
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 C0KE8

Quant Time: Jun 18 05:09:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR061518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Jun 18 04:02:12 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	544138	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	450903	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	136042	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	70265	3.79	ug/L	-0.02
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.80%
7) Chloroethane-d5	2.63	69	60973	4.48	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.60%
11) 1,1-Dichloroethene-d2	3.63	63	132241	2.77	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	55.40%#
20) 2-Butanone-d5	6.59	46	270447	54.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	108.88%
24) Chloroform-d	7.20	84	295768	4.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.40%
26) 1,2-Dichloroethane-d4	7.89	65	116674	4.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.00%
32) Benzene-d6	7.86	84	588770	3.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.00%
36) 1,2-Dichloropropane-d6	8.90	67	189618	4.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	87.40%
41) Toluene-d8	9.97	98	462692	3.42	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.40%#
43) trans-1,3-Dichloropropene-	10.23	79	21681	1.88	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	37.60%#
46) 2-Hexanone-d5	10.59	63	207164	51.64	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.28%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	89701	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	111473	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%


Target Compounds				Ovalue
13) Acetone	3.70	43	12951	5.772 ug/L 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA R\DATA\VR061618\
Data File : VR025352.D
Acq On : 17 Jun 2018 8:18
Operator : SY/MD
Sample : J3513-12
Misc : 25mL/MSVOA R/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0KE8

Quant Time: Jun 18 05:09:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR061518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Jun 18 04:02:12 2018
Response via : Initial Calibration





#13

Acetone

Concen: 5.772 ug/L

RT: 3.70 min Scan# 390

Delta R.T. 0.00 min

Lab File: VR025352.D

Acq: 17 Jun 2018 8:18

Instrument :
MSVOA_R
ClientSampleId :
C0KE8

Tot Ion: 43 Resp: 12951

Ion Ratio Lower Upper

43 100

58 16.5 0.0 41.6

