

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR070518\
 Data File : VR025447.D
 Acq On : 5 Jul 2018 19:42
 Operator : SY/MD
 Sample : J3834-04
 Misc : 25mL/MSVOA R/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_R
 ClientSampleId :
 VHBLK01

Quant Time: Jul 06 01:50:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Jul 05 15:29:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	663770	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	567308	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	185929	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.15	65	98235	5.37	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	107.40%
7) Chloroethane-d5	2.62	69	84935	5.45	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.00%
11) 1,1-Dichloroethene-d2	3.62	63	217713	3.93	ug/L	-0.02
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.60%
20) 2-Butanone-d5	6.60	46	303805	53.45	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.90%
24) Chloroform-d	7.20	84	422439	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
26) 1,2-Dichloroethane-d4	7.90	65	171298	4.94	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.80%
32) Benzene-d6	7.86	84	916035	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.60%
36) 1,2-Dichloropropane-d6	8.90	67	267791	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	9.97	98	800618	4.81	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.20%
43) trans-1,3-Dichloropropene-	10.24	79	61017	4.18	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	83.60%
46) 2-Hexanone-d5	10.59	63	271595	57.59	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	115.18%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	117039	4.75	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.00%
64) 1,2-Dichlorobenzene-d4	13.51	152	178835	5.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	116.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA R\DATA\VR070518\
Data File : VR025447.D
Acq On : 5 Jul 2018 19:42
Operator : SY/MD
Sample : J3834-04
Misc : 25mL/MSVOA R/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
VHBLK01

Quant Time: Jul 06 01:50:10 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Jul 05 15:29:58 2018
Response via : Initial Calibration

