

Data File : VR025456.D
Acq On : 12 Jul 2018 17:33
Operator : SY/MD
Sample : J3960-02MS
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0AF9MS

Quant Time: Jul 13 01:20:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 13 00:37:37 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	8.46	114	739788	5.00	ug/L	0.00
28) Chlorobenzene-d5	11.29	117	597524	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	13.22	152	201826	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	2.14	65	105223	5.16	ug/L	-0.01
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.20%
7) Chloroethane-d5	2.62	69	86728	4.99	ug/L	-0.01
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.80%
11) 1,1-Dichloroethene-d2	3.63	63	310743	5.03	ug/L	-0.01
Spiked Amount	5.000	Range	60 - 125	Recovery	=	100.60%
20) 2-Butanone-d5	6.59	46	305949	48.30	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.60%
24) Chloroform-d	7.20	84	471883	4.68	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.60%
26) 1,2-Dichloroethane-d4	7.89	65	178520	4.62	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.40%
32) Benzene-d6	7.86	84	1025635	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.60%
36) 1,2-Dichloropropane-d6	8.90	67	289510	4.99	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.80%
41) Toluene-d8	9.97	98	872457	4.97	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.40%
43) trans-1,3-Dichloropropene-	10.24	79	74879	4.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.40%
46) 2-Hexanone-d5	10.59	63	267153	53.78	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	107.56%
57) 1,1,2,2-Tetrachloroethane-	12.36	84	120839	4.66	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.20%
64) 1,2-Dichlorobenzene-d4	13.51	152	188867	5.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.40%

Target Compounds

					Qvalue
12) 1,1-Dichloroethene	3.65	96	109803	4.965 ug/L	96
19) 1,1-Dichloroethane	5.70	63	34299	0.363 ug/L #	90
22) cis-1,2-Dichloroethene	6.67	96	80808	1.600 ug/L #	90
33) Benzene	7.91	78	1051732	5.106 ug/L	100
34) Trichloroethene	8.71	95	411819	7.854 ug/L	99
42) Toluene	10.03	91	1064193	5.313 ug/L	99
51) Chlorobenzene	11.31	112	605846	4.968 ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VR025456.D
Acq On : 12 Jul 2018 17:33
Operator : SY/MD
Sample : J3960-02MS
Misc : 25mL/MSVOA_R/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_R
ClientSampleId :
C0AF9MS

Quant Time: Jul 13 01:20:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_R\METHODS\SOMRTR070518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Jul 13 00:37:37 2018
Response via : Initial Calibration

